



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: heta

Parameter: COD

SUPERVISOR REVIEW BY: heta

Run Number: LB72587

Reagent/Standard	Lot/Log #
COD Stock std, 1000ppm	WP34115
COD ICV-LCS std, 50ppm	WP35452
COD ICV-LCS std, 50ppm	WP35452
COD CCV std, 50ppm	WP35451
COD Digestion Vials Low Level 0-150Mg/L	W1906

Temp In (C): 150	Date In: 09/04/2014	Time In: 12:38
Temp Out (C): 150	Date Out: 09/04/2014	Time Out: 14:40

Intercept: -0.0932

Slope: 1.0176

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/04/2014	13:48
2	CAL2	10	1	Water	9.000	09/04/2014	13:49
3	CAL3	50	1	Water	52.000	09/04/2014	13:49
4	CAL4	100	1	Water	102.000	09/04/2014	13:50
5	CAL5	150	1	Water	152.000	09/04/2014	13:51



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: heta

Parameter: COD

SUPERVISOR REVIEW BY: heta

Run Number: LB72587

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	48.244	09/04/2014	13:52
2	ICB				1	Water	0.000	0.092	09/04/2014	13:52
3	CCV1	50			1	Water	53.000	52.175	09/04/2014	16:37
4	CCB1				1	Water	0.000	0.092	09/04/2014	16:37
5	LB72587BL				1	Water	0.000	0.092	09/04/2014	16:37
6	LB72587BS	50			1	Water	48.000	47.261	09/04/2014	16:38
7	F3726-01				1	Water	20.000	19.746	09/04/2014	16:38
8	F3726-01DUP				1	Water	21.000	20.728	09/04/2014	16:39
9	F3726-02	50			1	Water	65.000	63.967	09/04/2014	16:39
10	F3726-03	50			1	Water	64.000	62.985	09/04/2014	16:39
11	F3726-04				1	Water	22.000	21.711	09/04/2014	16:40
12	F3726-05				1	Water	0.000	0.092	09/04/2014	16:40
13	F3732-01				1	Water	1.000	1.074	09/04/2014	16:40
14	F3732-03				1	Water	0.000	0.092	09/04/2014	16:41
15	CCV2	50			1	Water	48.000	47.261	09/04/2014	16:41
16	CCB2				1	Water	0.000	0.092	09/04/2014	16:41
17	F3732-05				1	Water	0.000	0.092	09/04/2014	16:42
18	F3732-07				1	Water	0.000	0.092	09/04/2014	16:42
19	F3732-09				1	Water	22.000	21.711	09/04/2014	16:42
20	F3732-11				1	Water	8.000	7.953	09/04/2014	16:43
21	F3732-11DUP				1	Water	9.000	8.936	09/04/2014	16:43
22	F3732-12	50			1	Water	57.000	56.106	09/04/2014	16:43
23	F3732-13	50			1	Water	56.000	55.123	09/04/2014	16:44
24	F3732-17				1	Water	0.000	0.092	09/04/2014	16:44
25	F3732-19				1	Water	1.000	1.074	09/04/2014	16:44
26	F3732-21				1	Water	26.000	25.642	09/04/2014	16:45
27	CCV3	50			1	Water	52.000	51.192	09/04/2014	16:45
28	CCB3				1	Water	0.000	0.092	09/04/2014	16:45
29	F3741-03				1	Water	33.000	32.521	09/04/2014	16:46
30	CCV4	50			1	Water	49.000	48.244	09/04/2014	16:46
31	CCB4				1	Water	0.000	0.092	09/04/2014	16:46