



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB84438

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4		11/11/2016	10:00
2	H5524-08		1	52.0	125.6		11/11/2016	10:20
3	H5583-01		1	>100.0	>212.0		11/11/2016	10:40
4	H5602-02		1	>100.0	>212.0		11/11/2016	11:05
5	H5602-02DUP		1	>100.0	>212.0		11/11/2016	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33