



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB84898

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP51468
calibration std. phosphate 0.5 ppm	WP51467
calibration std. phosphate 0.3 ppm	WP51466
calibration std. 0 ppm	WP51463
phosphate ICV std.	WP51469
phosphate CCV std.	WP51470
calibration std. phosphate 0.1 ppm	WP51465
calibration std. phosphate 0.05 ppm	WP51464
Combined reagent	WP51458
Phenolphthalein indicator	WP48972
Sodium hydroxide, 1N	WP48973

Intercept: 0.0003

Slope: 0.6515

Regression: 0.999991

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/14/2016	11:37
2	CAL2	0.05	1	50	50	0.033	12/14/2016	11:37
3	CAL3	0.10	1	50	50	0.065	12/14/2016	11:38
4	CAL4	0.30	1	50	50	0.195	12/14/2016	11:38
5	CAL5	0.50	1	50	50	0.328	12/14/2016	11:39
6	CAL6	1.00	1	50	50	0.651	12/14/2016	11:39



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB84898

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.489	12/14/2016	11:40
2	ICB		1	50	50	0.000	0.000	12/14/2016	11:40
3	CCV1	0.50	1	50	50	0.326	0.500	12/14/2016	11:41
4	CCB1		1	50	50	0.000	0.000	12/14/2016	11:41
5	PB95325BL		1	50	50	0.000	0.000	12/14/2016	11:42
6	PB95325BS	0.50	1	50	50	0.319	0.489	12/14/2016	11:42
7	H5939-01		1	50	50	0.662	1.016	12/14/2016	11:43
8	H5939-01DUP		1	50	50	0.663	1.017	12/14/2016	11:43
9	H5939-01MS	0.50	1	50	50	0.702	1.077	12/14/2016	11:44
10	H5939-01MSD	0.50	1	50	50	0.735	1.128	12/14/2016	11:44
11	H5939-01		10	50	50	0.085	0.130	12/14/2016	11:45
12	H5939-01DUP		10	50	50	0.086	0.132	12/14/2016	11:45
13	H5939-01MS	0.50	10	50	50	0.115	0.176	12/14/2016	11:46
14	H5939-01MSD	0.50	10	50	50	0.114	0.175	12/14/2016	11:46
15	CCV2	0.50	1	50	50	0.327	0.501	12/14/2016	11:47
16	CCB2		1	50	50	0.000	0.000	12/14/2016	11:47