

LB85077

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85077
Instrument: Appolo-9000

M9060A - TOC-18

Analytical Summary Report

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85077
Instrument: Appolo-9000


Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb*	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	6221.000	1	11/29/16
2	0.5PPM	0.5PPM	CAL	32872.000	1	11/29/16
3	1.0PPM	1.0PPM	CAL	65081.000	1	11/29/16
4	2.0PPM	2.0PPM	CAL	95053.000	1	11/29/16
5	5.0PPM	5.0PPM	CAL	231731.000	1	11/29/16
6	10.0PPM	10.0PPM	CAL	415248.000	1	11/29/16
7	20.0PPM	20.0PPM	CAL	826802.000	1	11/29/16
8	ICV	ICV	ICV	10.092	1	11/29/16
9	ICB	ICB	ICB	0.000	1	11/29/16
10	CCV	CCV	CCV	10.023	1	12/28/16
11	CCB	CCB	CCB	0.000	1	12/28/16
12	LB85077BLW	LB85077BLW	MB	0.000	1	12/28/16
13	LB85077BSW	LB85077BSW	LCS	9.032	1	12/28/16
14	H6249-01	755015-MMW-03	SAM	9.241	2	12/28/16
15	H6249-02	755015-MMW-02	SAM	8.586	2	12/28/16
16	H6249-03S	755015-MMW-02S	MS	12.590	2	12/28/16
17	H6249-04SD	755015-MMW-02SD	MSD	12.720	2	12/28/16
18	H6277-01	755015-MW-14	SAM	4.338	1	12/28/16
19	H6277-02	755015-MMW-01	SAM	13.836	1	12/28/16
20	H6277-03	755015-TPM-01	SAM	17.366	2	12/28/16
21	H6277-04	755015-MMW-04	SAM	15.061	2	12/28/16
22	CCV	CCV	CCV	9.444	1	12/28/16
23	CCB	CCB	CCB	0.000	1	12/28/16
24	H6277-05	755015-TPMW-06	SAM	9.798	1	12/28/16
25	H6277-06	755015-TPM-02	SAM	12.714	2	12/28/16
26	H6277-07	755015-TPM-03	SAM	11.492	2	12/28/16
27	H6277-08	755015-TPMW-3	SAM	11.516	1	12/28/16
28	H6277-09	755015-TPMW-4	SAM	14.979	1	12/28/16
29	H6277-11	755015-DUPLICATE	SAM	14.816	1	12/28/16
30	CCV	CCV	CCV	10.374	1	12/28/16
31	CCB	CCB	CCB	0.000	1	12/28/16

* Cal results are in raw amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/3/17
Analyst : AK
Data File : LB85077.MDB

Approved By : 
Approved Date : 
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
TOC												
0.0PPM	0.0PPM				W	11/29/16						
TOC		PASS	6221.000		6220						mg/L	
0.5PPM	0.5PPM				W	11/29/16						
TOC		PASS	32872.000		32900						mg/L	
1.0PPM	1.0PPM				W	11/29/16						
TOC		PASS	65081.000		65100						mg/L	
2.0PPM	2.0PPM				W	11/29/16						
TOC		PASS	95053.000		95100						mg/L	
5.0PPM	5.0PPM				W	11/29/16						
TOC		PASS	231731.000		232000						mg/L	
10.0PPM	10.0PPM				W	11/29/16						
TOC		PASS	415248.000		415000						mg/L	
20.0PPM	20.0PPM				W	11/29/16						
TOC		PASS	826802.000		827000						mg/L	
ICV1	ICV1				W	11/29/16						
TOC		PASS	10.092		10.10	101.0	90	110			mg/L	
ICB1	ICB1				W	11/29/16						
TOC		PASS	0.000		0.000			+/-0.0970			mg/L	
CCV1	CCV1				W	12/28/16						
TOC		PASS	10.023		10.00	100.0	90	110			mg/L	
CCB1	CCB1				W	12/28/16						
TOC		PASS	0.000		0.000			+/-0.0970			mg/L	
LB85077BLW	LB85077BLW				W	12/28/16						
TOC		PASS	0.000		0.000			+/-0.0970			mg/L	
LB85077BSW	LB85077BSW				W	12/28/16						
TOC		PASS	9.032		9.03	90.0	90	110			mg/L	
H6249-01	755015-MMW-03	2			W	12/28/16						
TOC		PASS	9.241		18.5						mg/L	
H6249-02	755015-MMW-02	2			W	12/28/16						
TOC		PASS	8.586		17.2						mg/L	
H6249-03S	755015-MMW-02S	2			W	12/28/16						
TOC		PASS	12.590		25.20	80.0	75	125			mg/L	
H6249-04SD	755015-MMW-02SD	2			W	12/28/16						
TOC		PASS	12.720		25.40	82.0	75	125	0.8	20	mg/L	
H6277-01	755015-MW-14	1			W	12/28/16						
TOC		PASS	4.338		4.34						mg/L	
H6277-02	755015-MMW-01	1			W	12/28/16						
TOC		PASS	13.836		13.8						mg/L	
H6277-03	755015-TPM-01	2			W	12/28/16						
TOC		PASS	17.366		34.7						mg/L	

Cal results
are in raw
amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/3/17
Analyst : AK
Data File : LB85077.MDB

Approved By : 
Approved Date :
Worksheet # : 1/3/17

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
TOC												
H6277-04	755015-MMW-04			2	W	12/28/16						
TOC	PASS		15.061		30.1						mg/L	
CCV2	CCV2				W	12/28/16						
TOC	PASS		9.444		9.44	94.0	90	110			mg/L	
CCB2	CCB2				W	12/28/16						
TOC	PASS		0.000		0.000			+/-0.0970			mg/L	
H6277-05	755015-TPMW-06			1	W	12/28/16						
TOC	PASS		9.798		9.80						mg/L	
H6277-06	755015-TPM-02			2	W	12/28/16						
TOC	PASS		12.714		25.4						mg/L	
H6277-07	755015-TPM-03			2	W	12/28/16						
TOC	PASS		11.492		23.0						mg/L	
H6277-08	755015-TPMW-3			1	W	12/28/16						
TOC	PASS		11.516		11.5						mg/L	
H6277-09	755015-TPMW-4			1	W	12/28/16						
TOC	PASS		14.979		15.0						mg/L	
H6277-11	755015-DUPLICATE			1	W	12/28/16						
TOC	PASS		14.816		14.8						mg/L	
CCV3	CCV3				W	12/28/16						
TOC	PASS		10.374		10.40	104.0	90	110			mg/L	
CCB3	CCB3				W	12/28/16						
TOC	PASS		0.000		0.000			+/-0.0970			mg/L	