


ISM02.3

LB 85143
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40894



A= 0.0000e+000

B= 2.4419e-004

C= -1.0085e-002

Rho= 0.9999909

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.00	0.000	-0.012	-0.012	-8	0.000	-8					-
0.05	0.050	0.044	-0.006	220	0.0 %	220					-12
0.20	0.200	0.191	-0.009	823	0.0 %	823					-5
2.50	2.500	2.521	0.021	10366	0.0 %	10366					1
5.00	5.000	5.016	0.016	20582	0.0 %	20582					0
7.50	7.500	7.514	0.014	30814	0.0 %	30814					0
10.00	10.000	9.976	-0.024	40894	0.0 %	40894					0

NB
12/30/16

LB85143
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc Method	Units	Date	Type
	0 S0	-8	-	0 ISM02.3	PPB	12/30/2016 15:53	Std
	0.05 S0.05	220	-	0.05 ISM02.3	PPB	12/30/2016 15:55	Std
	0.2 S0.2	823	-	0.2 ISM02.3	PPB	12/30/2016 15:58	Std
	2.5 S2.5	10366	-	2.5 ISM02.3	PPB	12/30/2016 16:00	Std
	5 S5	20582	-	5 ISM02.3	PPB	12/30/2016 16:02	Std
	7.5 S7.5	30814	-	7.5 ISM02.3	PPB	12/30/2016 16:04	Std
	10 S10	40894	-	10 ISM02.3	PPB	12/30/2016 16:06	Std
ICV	ICV	15878	3.8672 -	ISM02.3	PPB	12/30/2016 16:10	SMPL
ICB	ICB	-17	-0.0142 -	ISM02.3	PPB	12/30/2016 16:12	SMPL
CCV007	CCV007	20429	4.9785 -	ISM02.3	PPB	12/30/2016 16:14	SMPL
CCB007	CCB007	-32	-0.0179 -	ISM02.3	PPB	12/30/2016 16:16	SMPL
PB95706BL	PBW003	-6	-0.0115 -	ISM02.3	PPB	12/30/2016 16:18	SMPL
H6284-01	MYA909	556	0.1257 -	ISM02.3	PPB	12/30/2016 16:21	SMPL
H6284-02	MYA909D	518	0.1164 -	ISM02.3	PPB	12/30/2016 16:23	SMPL
H6284-03	MYA909S	4995	1.2097 -	ISM02.3	PPB	12/30/2016 16:25	SMPL
H6284-04	MYA911	810	0.1877 -	ISM02.3	PPB	12/30/2016 16:27	SMPL
H6284-05	MYA919	1736	0.4138 -	ISM02.3	PPB	12/30/2016 16:29	SMPL
H6284-06	MYA935	905933	HIGH -	ISM02.3	PPB	12/30/2016 16:31	
H6284-07	MYA937	340679	HIGH -	ISM02.3	PPB	12/30/2016 16:33	
H6284-08	MYA939	125891	30.7316 -	ISM02.3	PPB	12/30/2016 16:35	SMPL
H6284-09	MYA943	1133095	HIGH -	ISM02.3	PPB	12/30/2016 16:37	
H6284-10	MYA945	199487	48.7032 -	ISM02.3	PPB	12/30/2016 16:39	SMPL
H6284-11	MYA947	634	0.1447 -	ISM02.3	PPB	12/30/2016 16:42	SMPL
H6284-12	MYA949	4442	1.0746 -	ISM02.3	PPB	12/30/2016 16:44	SMPL
H6284-13	MYA955	621532	HIGH -	ISM02.3	PPB	12/30/2016 16:46	
H6284-15	MYA963	244700	59.7439 -	ISM02.3	PPB	12/30/2016 16:50	SMPL
PB95707BL	PBW003	-7	-0.0118 -	ISM02.3	PPB	12/30/2016 16:54	SMPL
H6285-01	MYA910	297	0.0624 -	ISM02.3	PPB	12/30/2016 16:56	SMPL
H6285-02	MYA910D	328	0.07 -	ISM02.3	PPB	12/30/2016 17:01	SMPL
CCV008	CCV008	20726	5.0511 -	ISM02.3	PPB	12/30/2016 17:03	SMPL
CCB008	CCB008	-33	-0.0181 -	ISM02.3	PPB	12/30/2016 17:05	SMPL
H6285-03	MYA910S	4523	1.0944 -	ISM02.3	PPB	12/30/2016 17:07	SMPL
H6285-04	MYA912	341	0.0732 -	ISM02.3	PPB	12/30/2016 17:09	SMPL
H6285-05	MYA936	1735	0.4136 -	ISM02.3	PPB	12/30/2016 17:11	SMPL
H6285-06	MYA938	4626	1.1196 -	ISM02.3	PPB	12/30/2016 17:14	SMPL
H6285-07	MYA940	34079	8.3118 -	ISM02.3	PPB	12/30/2016 17:16	SMPL
H6285-08	MYA944	36558	8.9171 -	ISM02.3	PPB	12/30/2016 17:18	SMPL
H6285-09	MYA946	1004	0.2351 -	ISM02.3	PPB	12/30/2016 17:20	SMPL
H6285-10	MYA956	12912	3.1429 -	ISM02.3	PPB	12/30/2016 17:22	SMPL
H6285-11	MYA960	765	0.1767 -	ISM02.3	PPB	12/30/2016 17:24	SMPL
H6285-12	MYA964	4600	1.1132 -	ISM02.3	PPB	12/30/2016 17:26	SMPL
H6285-13	MYA966	1235	0.2915 -	ISM02.3	PPB	12/30/2016 17:28	SMPL
CCV009	CCV009	20628	5.0271 -	ISM02.3	PPB	12/30/2016 17:31	SMPL
CCB009	CCB009	-12	-0.013 -	ISM02.3	PPB	12/30/2016 17:33	SMPL
H6284-06DLX10	MYA935	187848	45.861 -	ISM02.3	PPB	12/30/2016 17:35	SMPL
H6284-07DLX10	MYA937	43435	10.5964 -	ISM02.3	PPB	12/30/2016 17:37	SMPL
H6284-08DLX10	MYA939	12743	3.1017 -	ISM02.3	PPB	12/30/2016 17:39	SMPL
H6284-09DLX10	MYA943	330593	HIGH -	ISM02.3	PPB	12/30/2016 17:41	
H6284-10DLX10	MYA945	20939	5.1031 -	ISM02.3	PPB	12/30/2016 17:43	SMPL
H6284-13DLX10	MYA955	103692	25.3108 -	ISM02.3	PPB	12/30/2016 17:45	SMPL
H6284-15DLX10	MYA963	24967	6.0867 -	ISM02.3	PPB	12/30/2016 17:47	SMPL
H6284-14	MYA959	880	0.2048 -	ISM02.3	PPB	12/30/2016 17:50	SMPL
H6284-16	MYA963 MYA965	707	0.1626 -	ISM02.3	PPB	12/30/2016 17:52	SMPL
CCV010	CCV010	20669	5.0371 -	ISM02.3	PPB	12/30/2016 17:54	SMPL
CCB010	CCB010	-40	-0.0199 -	ISM02.3	PPB	12/30/2016 17:56	SMPL
H6284-06DL2X100	MYA935	18898	4.6047 -	ISM02.3	PPB	12/30/2016 17:58	SMPL
H6284-07DL2X50	MYA937	8096	1.9669 -	ISM02.3	PPB	12/30/2016 18:00	SMPL
H6284-13DL2X50	MYA955	19622	4.7815 -	ISM02.3	PPB	12/30/2016 18:04	SMPL
H6284-09DL2X500	MYA943	8497	2.0648 -	ISM02.3	PPB	12/30/2016 18:07	SMPL
CCV011	CCV011	20478	4.9905 -	ISM02.3	PPB	12/30/2016 18:09	SMPL
CCB011	CCB011	-44	-0.0208 -	ISM02.3	PPB	12/30/2016 18:11	SMPL