



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB85150
Balance ID #: Gravimetric
Extraction IN Time: 01/03/2017 09:30

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 01/03/2017
Analysis Date: 01/03/2017
Extraction OUT Time: 01/03/2017 10:20

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight	Weight After Drying	Final Weight After Drying	Change Weight (g)	Result in ppm
1	LB85150BL	LB85150BL	WATER	2	1000	100	2.8547	2.8547	2.8547	2.8547	0	0
2	LB85150BS	LB85150BS	WATER	2	1000	100	2.9635	2.9635	2.9827	2.9827	0.0192	19.2
3	LB85150BSD	LB85150BSD	WATER	2	1000	100	2.0316	2.0316	2.051	2.051	0.0194	19.4
4	H6323-01	GRAB	WATER	2	1000	100	2.9805	2.9805	3.0611	3.0611	0.0806	80.6



QC Batch# LB85150 Test: TPH Analysis Date: 01/03/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2196
Magnesium Sulfate	
Sodium Sulfate	EP1711
1:1 HCL	WP50630
Silica Gel	W2061
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP50631
LCSWD	5.00 ML	WP50631
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID #

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 11:00
1.0000 gram Balance: 1.0004 In Time: 10:30
Bal Check Time: 09:40

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 12:30
1.0000 gram Balance: 1.0005 Out Time: 12:00
Bal Check Time: 13:10

WORKLIST(Hardcopy Internal Chain)

LB85150

WorkList Name : tph h6323

WorkList ID : 94054

Date : 1/3/2017 9:13:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	H6323-01	TPH	Conc H2SO4 to	ARAM01	O31	GRAB	12/30/2016	1664A

Date/Time 01.03.17 01:13 PM
 Received by: JP
 Relinquished by: JP

Date/Time 01.03.17 11:30 AM
 Received by: JP
 Relinquished by: JP