


7470A

LB85155
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39469



A= 0.0000e+000

B= 2.5044e-004

C= 2.3912e-002

Rho= 0.9996955

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	5	0.000	5				
0.2	0.200	0.218	0.018	773	0.0 %	773				
2.5	2.500	2.432	-0.068	9614	0.0 %	9614				
5.0	5.000	4.938	-0.062	19624	0.0 %	19624				
7.5	7.500	7.679	0.179	30566	0.0 %	30566				
10.0	10.000	9.908	-0.092	39469	0.0 %	39469				

LB85155 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jan 2017 14:02:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	5	-	0.0000	7470A	PPB	03 Jan 2017 14:07:12	S	Std
0.2 - 1	SO.2	773	-	0.2000	7470A	PPB	03 Jan 2017 14:09:20	S	Std
2.5 - 1	S2.5	9614	-	2.5000	7470A	PPB	03 Jan 2017 14:11:29	S	Std
5.0 - 1	S5	19624	-	5.0000	7470A	PPB	03 Jan 2017 14:13:37	S	Std
7.5 - 1	S7.5	30566	-	7.5000	7470A	PPB	03 Jan 2017 14:15:44	S	Std
10.0 - 1	S10	39469	-	10.0000	7470A	PPB	03 Jan 2017 14:17:51	S	Std
ICV83 - 1	ICV83	14998	3.7800	-	7470A	PPB	03 Jan 2017 14:20:35	U	SMPL
ICB83 - 1	ICB83	-14	0.0204	-	7470A	PPB	03 Jan 2017 14:22:41	U	SMPL
CCV12 - 1	CCV12	19343	4.8681	-	7470A	PPB	03 Jan 2017 14:24:48	U	SMPL
CCB12 - 1	CCB12	-19	0.0192	-	7470A	PPB	03 Jan 2017 14:26:55	U	SMPL
CRA - 1	CRA	735	0.2080	-	7470A	PPB	03 Jan 2017 14:29:01	U	SMPL
HighStd - 1	HighStd	38655	9.7046	-	7470A	PPB	03 Jan 2017 14:31:08	U	SMPL
ChkStd - 1	ChkStd	25013	6.2881	-	7470A	PPB	03 Jan 2017 14:33:15	U	SMPL
PB95708BL - 1	PBW	-40	0.0139	-	7470A	PPB	03 Jan 2017 14:35:22	U	SMPL
PB95708BS - 1	LCSW	13871	3.4977	-	7470A	PPB	03 Jan 2017 14:37:29	U	SMPL
PB95708TB - 1	PB95688TB	-25	0.0177	-	7470A	PPB	03 Jan 2017 14:39:37	U	SMPL
H6311-01 - 1	RPCXRAY01	7	0.0257	-	7470A	PPB	03 Jan 2017 14:41:46	U	SMPL
H6311-01DUP - 1	RPCXRAY01DUP	-5	0.0227	-	7470A	PPB	03 Jan 2017 14:43:54	U	SMPL
H6311-01MS - 1	RPCXRAY01MS	13704	3.4559	-	7470A	PPB	03 Jan 2017 14:46:02	U	SMPL
H6311-01MSD - 1	RPCXRAY01MSD	14106	3.5566	-	7470A	PPB	03 Jan 2017 14:48:09	U	SMPL
CCV13 - 1	CCV13	19297	4.8566	-	7470A	PPB	03 Jan 2017 14:50:15	U	SMPL
CCB13 - 1	CCB13	-17	0.0197	-	7470A	PPB	03 Jan 2017 14:52:22	U	SMPL
H6311-01LX5 - 1		16	0.0279	-	7470A	PPB	03 Jan 2017 14:54:29	U	SMPL
H6311-01A - 1		15013	3.7837	-	7470A	PPB	03 Jan 2017 14:56:35	U	SMPL
CCV14 - 1	CCV14	19678	4.9520	-	7470A	PPB	03 Jan 2017 14:58:42	U	SMPL
CCB14 - 1	CCB14	-39	0.0141	-	7470A	PPB	03 Jan 2017 15:00:49	U	SMPL