


7470A

LB85160
INSTRUMENT ID: CXL

Linear

 μ Abs.:

37447



A= 0.0000e+000

B= 2.6290e-004

C= 1.0946e-002

Rho= 0.9995921

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-18	0.000	-18				
0.2	0.200	0.188	-0.012	675	0.0 %	675				
2.5	2.500	2.457	-0.043	9303	0.0 %	9303				
5.0	5.000	4.984	-0.016	18917	0.0 %	18917				
7.5	7.500	7.709	0.209	29280	0.0 %	29280				
10.0	10.000	9.856	-0.144	37447	0.0 %	37447				

LB85160 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jan 2017 16:10:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.0000	7470A	PPB	03 Jan 2017 16:19:23	S	Std
0.2 - 1	50.2	675	-	0.2000	7470A	PPB	03 Jan 2017 16:21:31	S	Std
2.5 - 1	52.5	9303	-	2.5000	7470A	PPB	03 Jan 2017 16:23:39	S	Std
5.0 - 1	55	18917	-	5.0000	7470A	PPB	03 Jan 2017 16:25:48	S	Std
7.5 - 1	57.5	29280	-	7.5000	7470A	PPB	03 Jan 2017 16:27:55	S	Std
10.0 - 1	510	37447	-	10.0000	7470A	PPB	03 Jan 2017 16:32:30	S	Std
ICV84 - 1	ICV84	14810	3.9045	-	7470A	PPB	03 Jan 2017 16:35:10	U	SMPL
ICB84 - 1	ICB84	-34	0.0020	-	7470A	PPB	03 Jan 2017 16:37:15	U	SMPL
CCV15 - 1	CCV15	18586	4.8972	-	7470A	PPB	03 Jan 2017 16:39:22	U	SMPL
CCB15 - 1	CCB15	-25	0.0044	-	7470A	PPB	03 Jan 2017 16:41:29	U	SMPL
CRA - 1	CRA	666	0.1860	-	7470A	PPB	03 Jan 2017 16:43:35	U	SMPL
HighStd - 1	HighStd	37783	9.9441	-	7470A	PPB	03 Jan 2017 16:45:41	U	SMPL
ChkStd - 1	ChkStd	26166	6.8900	-	7470A	PPB	03 Jan 2017 16:47:48	U	SMPL
PB95731BL - 1	PBW	-58	-0.0043	-	7470A	PPB	03 Jan 2017 16:49:55	U	SMPL
PB95731BS - 1	LCSW	14565	3.8401	-	7470A	PPB	03 Jan 2017 16:52:02	U	SMPL
PB95731TB - 1	PB95709TB	-27	0.0038	-	7470A	PPB	03 Jan 2017 16:54:10	U	SMPL
H6322-04 - 1	TP9-COMP9	100	0.0372	-	7470A	PPB	03 Jan 2017 16:56:18	U	SMPL
H6322-04DUP - 1	TP9-COMP9DUP	82	0.0325	-	7470A	PPB	03 Jan 2017 16:58:26	U	SMPL
H6322-04MS - 1	TP9-COMP9MS	13751	3.6261	-	7470A	PPB	03 Jan 2017 17:00:35	U	SMPL
H6322-04MSD - 1	TP9-COMP9MSD	13917	3.6697	-	7470A	PPB	03 Jan 2017 17:02:41	U	SMPL
CCV16 - 1	CCV16	18809	4.9558	-	7470A	PPB	03 Jan 2017 17:04:48	U	SMPL
CCB16 - 1	CCB16	-45	-0.0009	-	7470A	PPB	03 Jan 2017 17:06:55	U	SMPL
H6325-02 - 1	LAR1020	42	0.0220	-	7470A	PPB	03 Jan 2017 17:09:01	U	SMPL
H6322-04LX5 - 1		14	0.0146	-	7470A	PPB	03 Jan 2017 17:11:08	U	SMPL
H6322-04A - 1		15502	4.0864	-	7470A	PPB	03 Jan 2017 17:13:15	U	SMPL
CCV17 - 1	CCV17	18690	4.9246	-	7470A	PPB	03 Jan 2017 17:15:21	U	SMPL
CCB17 - 1	CCB17	-12	0.0078	-	7470A	PPB	03 Jan 2017 17:17:29	U	SMPL