


7471A

L985167

INSTRUMENT ID: CV2

 μ Abs.:

45294



Linear

A= 0.0000e+000

B= 2.2042e-004

C= 8.5131e-003

Rho= 0.9999538

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	13	0.000	13				
0.2	0.200	0.181	-0.019	784	0.0 %	784				
2.5	2.500	2.539	0.039	11482	0.0 %	11482				
5.0	5.000	4.937	-0.063	22359	0.0 %	22359				
7.5	7.500	7.539	0.039	34162	0.0 %	34162				
10.0	10.000	9.992	-0.008	45294	0.0 %	45294				

LB85167 INSTRUMENJT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 04 Jan 2017 11:24:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	13	-	0.0000	7471A	PPB	04 Jan 2017 12:04:32	S	Std
0.2 - 1	S0.2	784	-	0.2000	7471A	PPB	04 Jan 2017 12:06:49	S	Std
2.5 - 1	S2.5	11482	-	2.5000	7471A	PPB	04 Jan 2017 12:09:07	S	Std
5.0 - 1	S5	22359	-	5.0000	7471A	PPB	04 Jan 2017 12:11:24	S	Std
7.5 - 1	S7.5	34162	-	7.5000	7471A	PPB	04 Jan 2017 12:13:42	S	Std
10.0 - 1	S10	45294	-	10.0000	7471A	PPB	04 Jan 2017 12:15:58	S	Std
ICV85 - 1	ICV85	17609	3.8899	-7471A	PPB	PPB	04 Jan 2017 12:18:39	U	SMPL
ICB85 - 1	ICB85	-48	-0.0021	-7471A	PPB	PPB	04 Jan 2017 12:20:55	U	SMPL
CCV18 - 1	CCV18	22260	4.9151	-7471A	PPB	PPB	04 Jan 2017 12:23:10	U	SMPL
CCB18 - 1	CCB18	-49	-0.0023	-7471A	PPB	PPB	04 Jan 2017 12:25:26	U	SMPL
CRA - 1	CRA	779	0.1802	-7471A	PPB	PPB	04 Jan 2017 12:27:42	U	SMPL
HighStd - 1	HighStd	44683	9.8577	-7471A	PPB	PPB	04 Jan 2017 12:29:58	U	SMPL
ChkStd - 1	ChkStd	31997	7.0614	-7471A	PPB	PPB	04 Jan 2017 12:32:13	U	SMPL
PB95732BL - 1	PBS	-29	0.0021	-7471A	PPB	PPB	04 Jan 2017 12:34:29	U	SMPL
PB95732BS - 1	LCSS	16381	3.6193	-7471A	PPB	PPB	04 Jan 2017 12:36:46	U	SMPL
H6322-03 - 1	TP9-COMP9	174292	38.4264	-7471A	PPB	PPB	04 Jan 2017 12:39:02	U	SMPL
H6322-03MS - 1	TP9-COMP9MS	206017	45.4194	-7471A	PPB	PPB	04 Jan 2017 12:44:11	U	SMPL
H6322-03MSD - 1	TP9-COMP9MSD	225051	49.6149	-7471A	PPB	PPB	04 Jan 2017 12:46:29	U	SMPL
CCV19 - 1	CCV19	22003	4.8585	-7471A	PPB	PPB	04 Jan 2017 12:48:47	U	SMPL
CCB19 - 1	CCB19	-76	-0.0082	-7471A	PPB	PPB	04 Jan 2017 12:51:05	U	SMPL
H6322-03DUP - 1	TP9-COMP9DUP	175248	38.6372	-7471A	PPB	PPB	04 Jan 2017 12:53:22	U	SMPL
H6322-03LX5 - 1		34033	7.5102	-7471A	PPB	PPB	04 Jan 2017 12:55:38	U	SMPL
H6322-03A - 1		195740	43.1541	-7471A	PPB	PPB	04 Jan 2017 13:00:43	U	SMPL
H6322-03DLX10 - 1		17840	3.9409	-7471A	PPB	PPB	04 Jan 2017 13:02:58	U	SMPL
H6322-03DUPDLX10 - 1		17513	3.8688	-7471A	PPB	PPB	04 Jan 2017 13:05:17	U	SMPL
H6322-03MSDLX10 - 1		20509	4.5292	-7471A	PPB	PPB	04 Jan 2017 13:10:25	U	SMPL
H6322-03MSDDLX10 - 1		21052	4.6489	-7471A	PPB	PPB	04 Jan 2017 13:12:42	U	SMPL
H6322-03LDLX50 - 1		3283	0.7322	-7471A	PPB	PPB	04 Jan 2017 13:14:59	U	SMPL
H6322-03ADLX10 - 1		19657	4.3414	-7471A	PPB	PPB	04 Jan 2017 13:17:16	U	SMPL
CCV20 - 1	CCV20	22296	4.9231	-7471A	PPB	PPB	04 Jan 2017 13:26:22	U	SMPL
CCB20 - 1	CCB20	-21	0.0039	-7471A	PPB	PPB	04 Jan 2017 13:28:38	U	SMPL