

File Utility Help

RN↓ RN↑ ?

INSTRUMENT ID: CV2

Protocol chemtech

Dataset/Proto LB85171 /chemtech

Protocol Line info Cal Curve Report Ctrl Chart Viewer

Reset

Calib Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B

C

Rho

Type

Linear

Print Cal

☒ Calibrated☒ Accepted


Accept

 μ Abs.

49340

Accepted

New



04-Jan-17 14:42

Conc.

10.0

S	Conc.	Calc.	Dev.	Mean	SD or %R...	Rep 1	Rep 2	Rep 3	Rep 4	F
01	.00000	-.025	-.025	119	1	118				
02	.20000	.175	-.025	1102	0%	1102				
03	2.5000	2.56	.059	12877	0%	12877				
04	5.0000	4.98	-.022	24827	0%	24827				
05	7.5000	7.57	.073	37643	0%	37643				
06	10.0000	9.94	-.059	49340	0%	49340				

Ready

CAP NUM

Protocol: chemtech

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: HIGH STD								
					Seq: 11	14:58:05	04 Jan 17	HG
				HIGH STD				
Hg	9.92	PPB	49252					
=====								
*** Sample ID: CHK STD								
					Seq: 12	15:00:16	04 Jan 17	HG
				CHK STD				
Hg	7.07	PPB	35180					
=====								
*** Sample ID: PBW								
					Seq: 13	15:02:25	04 Jan 17	HG
				PB95748BL				
Hg	-.073	PPB	-123					
=====								
*** Sample ID: LCSW								
					Seq: 14	15:04:27	04 Jan 17	HG
				PB95748BS				
Hg	3.61	PPB	18055					
=====								
*** Sample ID: H6103-07								
					Seq: 15	15:07:02	04 Jan 17	HG
				LOD-WATER-07-QT1-2017				
Hg	.053	PPB	500					
=====								
*** Sample ID: H6103-08								
					Seq: 16	15:09:04	04 Jan 17	HG
				LOQ-WATER-08-QT1-2017				
Hg	.363	PPB	2032					
=====								
*** Sample ID: CCV128								
					Seq: 17	15:11:32	04 Jan 17	HG
				CCV128				
Hg	4.97	PPB	24789					
=====								
*** Sample ID: CCB128								
					Seq: 18	15:13:41	04 Jan 17	HG
				CCB128				
Hg	-.060	PPB	-57					
=====								