

Protocol chemtech

Dataset/Proto IB85172 /chemtech

Protocol | Line info | Cal Curve | Report | Ctrl Chart | Viewer

Reset

Calib Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B

C

Rho

Type

Linear

Print Cal

☒ Calibrated☒ Accepted


Accept

 μ Abs.

48903

Accepted

New



04-Jan-17 16:30

Conc.

10.0

S	Conc.	Calc.	Dev.	Mean	SD or %R...	Rep 1	Rep 2	Rep 3	Rep 4	F
01	.00000	-.001	-.001	-48	1	-48				
02	.20000	.218	.018	1049	0%	1048				
03	2.5000	2.37	-.134	11775	0%	11774				
04	5.0000	5.07	.070	25282	0%	25282				
05	7.5000	7.75	.247	38650	0%	38649				
06	10.000	9.80	-.200	48904	0%	48903				

Ready

CAP NUM

Protocol: chemtech

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: HighStd								
				Seq: 18	16:44:22	04 Jan 17	HG	
			HighStd					
Hg	9.87	PPB	49273					
=====								
*** Sample ID: ChkStd								
				Seq: 19	16:46:32	04 Jan 17	HG	
			ChkStd					
Hg	7.08	PPB	35306					
=====								
*** Sample ID: PBW								
				Seq: 20	16:49:10	04 Jan 17	HG	
			PB95749BL					
Hg	.008	PPB	0					
=====								
*** Sample ID: LCSW								
				Seq: 21	16:51:49	04 Jan 17	HG	
			PB95749BS					
Hg	3.94	PPB	19626					
=====								
*** Sample ID: H6103-05								
				Seq: 22	16:54:00	04 Jan 17	HG	
			LOD-WATER-05-QT1-2017					
Hg	.106	PPB	489					
=====								
*** Sample ID: H6103-06								
				Seq: 23	16:56:10	04 Jan 17	HG	
			LOQ-WATER-06-QT1-2017					
Hg	.437	PPB	2144					
=====								
*** Sample ID: CCV130								
				Seq: 24	16:58:40	04 Jan 17	HG	
			CCV130					
Hg	5.03	PPB	25082					
=====								
*** Sample ID: CCB130								
				Seq: 25	17:00:40	04 Jan 17	HG	
			CCB130					
Hg	-.017	PPB	-125					
=====								