



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85176

Temp.: 21.3 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2033
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/05/2017	08:45
2	CAL2	1	Water			7.00	01/05/2017	08:46
3	CAL3	1	Water			10.03	01/05/2017	08:47
4	ICV	1	Water			7.02	01/05/2017	08:48
5	CCV1	1	Water			2.02	01/05/2017	08:49
6	I1025-04	1	Solid	20.03	20	9.67	01/05/2017	08:50
7	I1025-04DUP	1	Solid	20.01	20	9.68	01/05/2017	08:51
8	CCV2	1	Water			12.02	01/05/2017	08:52

LB 85176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY i1025 WorkList ID : 94091 Date : 1/5/2017 8:02:38 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1025-04	Corrosivity	Cool 4 deg C	DEWB01	D32	TP7-8-COMP10	01/04/2017	9045C

Date/Time 01-05-17 8:02 AM
 Received by: JP
 Relinquished by: SR

Date/Time 01-05-17 9:30 AM
 Received by: SR
 Relinquished by: JP