



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85193

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP51900
COD CCV std, 50ppm	WP51901

Temp In (C): <u>150</u>	Date In: <u>01/05/2017</u>	Time In: <u>09:50</u>
Temp Out (C): <u>150</u>	Date Out: <u>01/05/2017</u>	Time Out: <u>11:53</u>

Intercept: 0.2469

Slope: 0.9992

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	12/30/2016	12:50
2	CAL2	10	1	Water	10.000	12/30/2016	12:50
3	CAL3	50	1	Water	51.000	12/30/2016	12:51
4	CAL4	100	1	Water	100.000	12/30/2016	12:51
5	CAL5	150	1	Water	150.000	12/30/2016	12:52

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85193

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.793	12/30/2016	12:52
2	ICB				1	Water	0.000	-0.247	12/30/2016	12:53
3	CCV1	50			1	Water	51.000	50.794	01/05/2017	13:09
4	CCB1				1	Water	0.000	-0.247	01/05/2017	13:09
5	LB85193BL				1	Water	0.000	-0.247	01/05/2017	13:10
6	LB85193BS	50			1	Water	53.000	52.795	01/05/2017	13:10
7	I1003-01				1	Water	5.000	4.757	01/05/2017	13:11
8	I1003-01DUP				1	Water	5.000	4.757	01/05/2017	13:11
9	I1003-01MS	50			1	Water	54.000	53.796	01/05/2017	13:12
10	I1003-01MSD	50			1	Water	55.000	54.797	01/05/2017	13:12
11	CCV2	50			1	Water	52.000	51.795	01/05/2017	13:13
12	CCB2				1	Water	0.000	-0.247	01/05/2017	13:13

2385193

WORKLIST(Hardcopy Internal Chain)

Worklist Name : COD010517

Worklist ID : 94132

Date : 1/5/2017 09:25:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1003-01	COD	Conc H2SO4 tk	NAT104	O33	DSN001	01/03/2017	SM5220 D

Date/Time 01/05/17 09:30
Received by: TB
Relinquished by: CP

Date/Time 01/05/17 / 12:00
Received by: CP
Relinquished by: TB