


245.1

LB 85198
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36860



A= 0.0000e+000

B= 2.7006e-004

C= 6.6043e-003

Rho= 0.9996387

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-15	0.000	-15				
0.2	0.200	0.208	0.008	747	0.0 %	747				
2.5	2.500	2.567	0.067	9480	0.0 %	9480				
5.0	5.000	4.819	-0.181	17820	0.0 %	17820				
7.5	7.500	7.642	0.142	28274	0.0 %	28274				
10.0	10.000	9.961	-0.039	36860	0.0 %	36860				

LB85198 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Jan 2017 14:24:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-15	-	0.0000	245.1	PPB	05 Jan 2017 14:40:57	S	Std
0.2 - 1	SO.2	747	-	0.2000	245.1	PPB	05 Jan 2017 14:43:06	S	Std
2.5 - 1	S2.5	9480	-	2.5000	245.1	PPB	05 Jan 2017 14:45:14	S	Std
5.0 - 1	S5	17820	-	5.0000	245.1	PPB	05 Jan 2017 14:47:22	S	Std
7.5 - 1	S7.5	28274	-	7.5000	245.1	PPB	05 Jan 2017 14:49:30	S	Std
10.0 - 1	S10	36860	-	10.0000	245.1	PPB	05 Jan 2017 14:51:37	S	Std
ICV86 - 1	ICV86	14849	4.0167	-245.1		PPB	05 Jan 2017 14:55:16	U	SMPL
ICB86 - 1	ICB86	-35	-0.0028	-245.1		PPB	05 Jan 2017 14:57:22	U	SMPL
CCV21 - 1	CCV21	17623	4.7659	-245.1		PPB	05 Jan 2017 14:59:29	U	SMPL
CCB21 - 1	CCB21	-31	-0.0018	-245.1		PPB	05 Jan 2017 15:01:36	U	SMPL
CRA - 1	CRA	726	0.2027	-245.1		PPB	05 Jan 2017 15:03:42	U	SMPL
HighStd - 1	HighStd	36131	9.7641	-245.1		PPB	05 Jan 2017 15:05:49	U	SMPL
ChkStd - 1	ChkStd	23684	6.4027	-245.1		PPB	05 Jan 2017 15:07:56	U	SMPL
PB95768BL - 1	PBW	-40	-0.0042	-245.1		PPB	05 Jan 2017 15:10:03	U	SMPL
PB95768BS - 1	LCSW	13774	3.7264	-245.1		PPB	05 Jan 2017 15:12:11	U	SMPL
I1002-08 - 1	MDL-WATER-1	168	0.0520	-245.1		PPB	05 Jan 2017 15:14:18	U	SMPL
I1002-09 - 1	MDL-WATER-2	189	0.0576	-245.1		PPB	05 Jan 2017 15:16:26	U	SMPL
I1002-10 - 1	MDL-WATER-3	177	0.0544	-245.1		PPB	05 Jan 2017 15:18:35	U	SMPL
I1002-11 - 1	MDL-WATER-4	176	0.0541	-245.1		PPB	05 Jan 2017 15:20:44	U	SMPL
I1002-12 - 1	MDL-WATER-5	163	0.0506	-245.1		PPB	05 Jan 2017 15:25:19	U	SMPL
CCV22 - 1	CCV22	17481	4.7275	-245.1		PPB	05 Jan 2017 15:27:25	U	SMPL
CCB22 - 1	CCB22	-15	0.0026	-245.1		PPB	05 Jan 2017 15:29:32	U	SMPL
I1002-13 - 1	MDL-WATER-6	144	0.0455	-245.1		PPB	05 Jan 2017 15:36:26	U	SMPL
I1002-14 - 1	MDL-WATER-7	151	0.0474	-245.1		PPB	05 Jan 2017 15:38:32	U	SMPL
CCV23 - 1	CCV23	17560	4.7489	-245.1		PPB	05 Jan 2017 15:40:39	U	SMPL
CCB23 - 1	CCB23	-24	0.0001	-245.1		PPB	05 Jan 2017 15:42:46	U	SMPL