


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85206.csv

Approved By : 
 Approved Date :
 Worksheet # : 1/2/17

1000.0 Inorganic Analysis - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 2
Bromide											
STD1	STD1				W	12/19/16					
Bromide		PASS	0.000		0.000					mg/L	
STD2	STD2				W	12/19/16					
Bromide		PASS	0.982		0.982					mg/L	
STD3	STD3				W	12/19/16					
Bromide		PASS	2.326		2.33					mg/L	
STD4	STD4				W	12/19/16					
Bromide		PASS	4.181		4.18					mg/L	
STD5	STD5				W	12/19/16					
Bromide		PASS	9.734		9.73					mg/L	
STD6	STD6				W	12/19/16					
Bromide		PASS	19.435		19.4					mg/L	
STD7	STD7				W	12/19/16					
Bromide		PASS	24.257		24.3					mg/L	
STD8	STD8				W	12/19/16					
Bromide		PASS	50.585		50.6					mg/L	
ICV1	ICV1				W	12/19/16					
Bromide		PASS	24.524		24.50	98.0	90	110		mg/L	
ICB1	ICB1				W	12/19/16					
Bromide		PASS	0.000		0.000			+/-0.0660		mg/L	
CCV1	CCV1				W	1/4/17					
Bromide		PASS	24.760		24.80	99.0	90	110		mg/L	
CCB1	CCB1				W	1/4/17					
Bromide		PASS	0.000		0.000			+/-0.0660		mg/L	
LB85206BLW	LB85206BLW				W	1/4/17					
Bromide		PASS	0.000		0.000			+/-0.0660		mg/L	
LB85206BSW	LB85206BSW				W	1/4/17					
Bromide		PASS	24.416		24.40	98.0	90	110		mg/L	
I1002-08	MDL-WATER-1			1	W	1/4/17					
Bromide		PASS	0.433		0.433					mg/L	
I1002-09	MDL-WATER-2			1	W	1/4/17					
Bromide		PASS	0.422		0.422					mg/L	
I1002-10	MDL-WATER-3			1	W	1/4/17					
Bromide		PASS	0.409		0.409					mg/L	
I1002-11	MDL-WATER-4			1	W	1/4/17					
Bromide		PASS	0.403		0.403					mg/L	
I1002-12	MDL-WATER-5			1	W	1/4/17					
Bromide		PASS	0.407		0.407					mg/L	
I1002-13	MDL-WATER-6			1	W	1/4/17					
Bromide		PASS	0.395		0.395					mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AP
Data File : LB85206.csv


Approved By : 
Approved Date :
Worksheet # : 1/2/17

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Bromide	PASS	0.422		0.422						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Bromide	PASS	0.436		0.436						mg/L	
CCV2	CCV2			W	1/4/17						
Bromide	PASS	24.790		24.80	99.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Bromide	PASS	0.000		0.000			+/-0.0660			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85206.csv

Approved By : 
 Approved Date : 1/4/18
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Chloride											
STD1	STD1				W	12/19/16					
Chloride		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/19/16					
Chloride		PASS	0.383		0.383						mg/L
STD3	STD3				W	12/19/16					
Chloride		PASS	0.770		0.770						mg/L
STD4	STD4				W	12/19/16					
Chloride		PASS	1.266		1.27						mg/L
STD5	STD5				W	12/19/16					
Chloride		PASS	2.897		2.90						mg/L
STD6	STD6				W	12/19/16					
Chloride		PASS	5.700		5.70						mg/L
STD7	STD7				W	12/19/16					
Chloride		PASS	7.167		7.17						mg/L
STD8	STD8				W	12/19/16					
Chloride		PASS	15.268		15.3						mg/L
ICV1	ICV1				W	12/19/16					
Chloride		PASS	7.282		7.28	97.0	90	110			mg/L
ICB1	ICB1				W	12/19/16					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
CCV1	CCV1				W	1/4/17					
Chloride		PASS	7.213		7.21	96.0	90	110			mg/L
CCB1	CCB1				W	1/4/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85206BLW	LB85206BLW				W	1/4/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85206BSW	LB85206BSW				W	1/4/17					
Chloride		PASS	7.141		7.14	95.0	90	110			mg/L
I1002-08	MDL-WATER-1			1	W	1/4/17					
Chloride		PASS	0.061		0.061						mg/L
I1002-09	MDL-WATER-2			1	W	1/4/17					
Chloride		PASS	0.052		0.052						mg/L
I1002-10	MDL-WATER-3			1	W	1/4/17					
Chloride		PASS	0.050		0.050						mg/L
I1002-11	MDL-WATER-4			1	W	1/4/17					
Chloride		PASS	0.060		0.060						mg/L
I1002-12	MDL-WATER-5			1	W	1/4/17					
Chloride		PASS	0.053		0.053						mg/L
I1002-13	MDL-WATER-6			1	W	1/4/17					
Chloride		PASS	0.052		0.052						mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AP
Data File : LB85206.csv

Approved By : 
Approved Date : 1/6/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Chloride	PASS	0.056		0.056						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Chloride	PASS	0.062		0.062						mg/L	
CCV2	CCV2			W	1/4/17						
Chloride	PASS	7.263		7.26	97.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Chloride	PASS	0.000		0.000			+/-0.0750			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85206.csv

Approved By : 
 Approved Date : 1/6/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Fluoride											
STD1	STD1				W	12/19/16					
Fluoride		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/19/16					
Fluoride		PASS	0.308		0.308						mg/L
STD3	STD3				W	12/19/16					
Fluoride		PASS	0.491		0.491						mg/L
STD4	STD4				W	12/19/16					
Fluoride		PASS	0.847		0.847						mg/L
STD5	STD5				W	12/19/16					
Fluoride		PASS	1.940		1.94						mg/L
STD6	STD6				W	12/19/16					
Fluoride		PASS	3.885		3.88						mg/L
STD7	STD7				W	12/19/16					
Fluoride		PASS	4.615		4.62						mg/L
STD8	STD8				W	12/19/16					
Fluoride		PASS	10.215		10.2						mg/L
ICV1	ICV1				W	12/19/16					
Fluoride		PASS	4.649		4.65	93.0	90	110			mg/L
ICB1	ICB1				W	12/19/16					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
CCV1	CCV1				W	1/4/17					
Fluoride		PASS	5.056		5.06	101.0	90	110			mg/L
CCB1	CCB1				W	1/4/17					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
LB85206BLW	LB85206BLW				W	1/4/17					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
LB85206BSW	LB85206BSW				W	1/4/17					
Fluoride		PASS	4.852		4.85	97.0	90	110			mg/L
I1002-08	MDL-WATER-1			1	W	1/4/17					
Fluoride		PASS	0.087		0.087						mg/L
I1002-09	MDL-WATER-2			1	W	1/4/17					
Fluoride		PASS	0.086		0.086						mg/L
I1002-10	MDL-WATER-3			1	W	1/4/17					
Fluoride		PASS	0.103		0.103						mg/L
I1002-11	MDL-WATER-4			1	W	1/4/17					
Fluoride		PASS	0.083		0.083						mg/L
I1002-12	MDL-WATER-5			1	W	1/4/17					
Fluoride		PASS	0.093		0.093						mg/L
I1002-13	MDL-WATER-6			1	W	1/4/17					
Fluoride		PASS	0.089		0.089						mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AP
Data File : LB85206.csv

Approved By : 
Approved Date : 1/8/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Fluoride											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Fluoride	PASS	0.084		0.084						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Fluoride	PASS	0.107		0.107						mg/L	
CCV2	CCV2			W	1/4/17						
Fluoride	PASS	5.155		5.16	103.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Fluoride	PASS	0.000		0.000			+/-0.0430			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85206.csv

Approved By : 
 Approved Date : 1/6/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Nitrate											
STD1	STD1				W	12/19/16					
Nitrate		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/19/16					
Nitrate		PASS	0.331		0.331						mg/L
STD3	STD3				W	12/19/16					
Nitrate		PASS	0.616		0.616						mg/L
STD4	STD4				W	12/19/16					
Nitrate		PASS	0.971		0.971						mg/L
STD5	STD5				W	12/19/16					
Nitrate		PASS	2.142		2.14						mg/L
STD6	STD6				W	12/19/16					
Nitrate		PASS	4.251		4.25						mg/L
STD7	STD7				W	12/19/16					
Nitrate		PASS	5.347		5.35						mg/L
STD8	STD8				W	12/19/16					
Nitrate		PASS	11.541		11.5						mg/L
ICV1	ICV1				W	12/19/16					
Nitrate		PASS	6.051		6.05	97.0	90	110			mg/L
ICB1	ICB1				W	12/19/16					
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L
CCV1	CCV1				W	1/4/17					
Nitrate		PASS	5.448		5.45	96.0	90	110			mg/L
CCB1	CCB1				W	1/4/17					
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L
LB85206BLW	LB85206BLW				W	1/4/17					
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L
LB85206BSW	LB85206BSW				W	1/4/17					
Nitrate		PASS	6.074		6.07	97.0	90	110			mg/L
I1002-08	MDL-WATER-1			1	W	1/4/17					
Nitrate		PASS	0.077		0.077						mg/L
I1002-09	MDL-WATER-2			1	W	1/4/17					
Nitrate		PASS	0.075		0.075						mg/L
I1002-10	MDL-WATER-3			1	W	1/4/17					
Nitrate		PASS	0.076		0.076						mg/L
I1002-11	MDL-WATER-4			1	W	1/4/17					
Nitrate		PASS	0.080		0.080						mg/L
I1002-12	MDL-WATER-5			1	W	1/4/17					
Nitrate		PASS	0.077		0.077						mg/L
I1002-13	MDL-WATER-6			1	W	1/4/17					
Nitrate		PASS	0.074		0.074						mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AP
Data File : LB85206.csv

Approved By : 
Approved Date : 1/5/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrate											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Nitrate	PASS	0.079		0.079						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Nitrate	PASS	0.078		0.078						mg/L	
CCV2	CCV2			W	1/4/17						
Nitrate	PASS	5.460		5.46	97.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Nitrate	PASS	0.000		0.000			+/-0.0270			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85206.csv


Approved By : 
 Approved Date : 1/4/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Sulfate										
STD1	STD1				W	12/19/16				
Sulfate		PASS	0.000		0.000				mg/L	
STD2	STD2				W	12/19/16				
Sulfate		PASS	2.472		2.47				mg/L	
STD3	STD3				W	12/19/16				
Sulfate		PASS	4.201		4.20				mg/L	
STD4	STD4				W	12/19/16				
Sulfate		PASS	6.489		6.49				mg/L	
STD5	STD5				W	12/19/16				
Sulfate		PASS	14.075		14.1				mg/L	
STD6	STD6				W	12/19/16				
Sulfate		PASS	27.937		27.9				mg/L	
STD7	STD7				W	12/19/16				
Sulfate		PASS	35.311		35.3				mg/L	
STD8	STD8				W	12/19/16				
Sulfate		PASS	76.765		76.8				mg/L	
ICV1	ICV1				W	12/19/16				
Sulfate		PASS	36.264		36.30	97.0	90	110	mg/L	
ICB1	ICB1				W	12/19/16				
Sulfate		PASS	0.000		0.000			+/-0.1320	mg/L	
CCV1	CCV1				W	1/4/17				
Sulfate		PASS	36.129		36.10	96.0	90	110	mg/L	
CCB1	CCB1				W	1/4/17				
Sulfate		PASS	0.000		0.000			+/-0.1320	mg/L	
LB85206BLW	LB85206BLW				W	1/4/17				
Sulfate		PASS	0.000		0.000			+/-0.1320	mg/L	
LB85206BSW	LB85206BSW				W	1/4/17				
Sulfate		PASS	35.447		35.40	94.0	90	110	mg/L	
I1002-08	MDL-WATER-1			1	W	1/4/17				
Sulfate		PASS	0.641		0.641				mg/L	
I1002-09	MDL-WATER-2			1	W	1/4/17				
Sulfate		PASS	0.679		0.679				mg/L	
I1002-10	MDL-WATER-3			1	W	1/4/17				
Sulfate		PASS	0.645		0.645				mg/L	
I1002-11	MDL-WATER-4			1	W	1/4/17				
Sulfate		PASS	0.662		0.662				mg/L	
I1002-12	MDL-WATER-5			1	W	1/4/17				
Sulfate		PASS	0.662		0.662				mg/L	
I1002-13	MDL-WATER-6			1	W	1/4/17				
Sulfate		PASS	0.655		0.655				mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AP
Data File : LB85206.csv

Approved By : 
Approved Date :
Worksheet # : 1/6/17

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Sulfate											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Sulfate	PASS	0.667		0.667						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Sulfate	PASS	0.634		0.634						mg/L	
CCV2	CCV2			W	1/4/17						
Sulfate	PASS	36.357		36.40	97.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Sulfate	PASS	0.000		0.000			+/-0.1320			mg/L	