

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/5/17

Analyst : AK

Data File : LB85210.csv

Approved By : 

Approved Date : 1/6/17

Worksheet # :

m9c12 A-B total Amenable of Reactive Cyanide - 16

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method				Line 1
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/5/17						
Cyanide		PASS	-1.061		-0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/5/17						
Cyanide		PASS	4.178		0.004						mg/L	
10PPBCN	10PPBCN				W	1/5/17						
Cyanide		PASS	9.015		0.009						mg/L	
50PPBCN	50PPBCN				W	1/5/17						
Cyanide		PASS	50.736		0.051						mg/L	
100PPBCN	100PPBCN				W	1/5/17						
Cyanide		PASS	102.328		0.102						mg/L	
250PPBCN	250PPBCN				W	1/5/17						
Cyanide		PASS	250.631		0.251						mg/L	
500PPBCN	500PPBCN				W	1/5/17						
Cyanide		PASS	499.173		0.499						mg/L	
ICV1	ICV1				W	1/5/17						
Cyanide		PASS	96.129		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/5/17						
Cyanide		PASS	-0.243		0.000			+/-0.0010			mg/L	
CCV1	CCV1				W	1/5/17						
Cyanide		PASS	241.392		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/5/17						
Cyanide		PASS	0.117		0.000			+/-0.0010			mg/L	
MIDPB95774	MIDPB95774				W	1/5/17						
Cyanide		PASS	236.380		0.236						mg/L	
PB95774BLW	PB95774BLW				W	1/5/17						
Cyanide		PASS	-1.234		-0.001			+/-0.0010			mg/L	
PB95774BSW	PB95774BSW				W	1/5/17						
Cyanide		PASS	98.310		0.098	98.0	85	115			mg/L	
I1027-01	EFF-WW			1	W	1/5/17						
Cyanide		PASS	-0.915		-0.001						mg/L	
I1027-01D	EFF-WWD			1	W	1/5/17						
Cyanide		PASS	-0.427		0.000				0	20	mg/L	
I1027-01S	EFF-WWS			1	W	1/5/17						
Cyanide		PASS	36.368		0.036	90.0	75	125			mg/L	
I1027-01SD	EFF-WWSD			1	W	1/5/17						
Cyanide		PASS	36.327		0.036	90.0	75	125	0	20	mg/L	
CCV2	CCV2				W	1/5/17						
Cyanide		PASS	243.167		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	1/5/17						
Cyanide		PASS	-0.576		-0.001			+/-0.0010			mg/L	