

ad
1/6/17

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
BLANK	51902			TOC	
250mg/l	836527			TOC	
500mg/l	1562927			TOC	
1000mg/l.....	3200398...	..	...	TOC	..
2000mg/l	6050279			TOC	
ICV	949.0182			TOC	
ICB.....	8.9775...	..	...	TOC	..
CCV	966.0576			TOC	
CCB	71.8154			TOC	
LB85213BLS.....	8.0863...	..	...	TOC	..
LB85213BLS	14.2636			TOC	
LB85213BLS	10.2379			TOC	
LB85213BLS.....	9.1038...	..	...	TOC	..
LB85213BSS	956.8993			TOC	
LB85213BSS	1013.8410			TOC	
LB85213BSS.....	970.1638...	..	...	TOC	..
LB85213BSS	915.0014			TOC	
H6103-01	52.6161			TOC	
H6103-01.....	59.1570...	..	...	TOC	..
H6103-01	51.1804			TOC	
H6103-01	53.8688			TOC	
H6103-02.....	553.5618...	..	...	TOC	..
H6103-02	585.5286			TOC	
H6103-02	546.5222			TOC	
H6103-02.....	602.6577...	..	...	TOC	..
CCV	985.2183			TOC	
CCB	36.0390			TOC	

MS
1/6/17

Method ID	Sample Type	Vial Timestamp	Message
Boat Sampler	TOC Standard	2017/01/06 10:24	
Boat Sampler	TOC Standard	2017/01/06 10:28	
Boat Sampler	TOC Standard	2017/01/06 10:33	
Boat Sampler	...TOC Standard	.. 2017/01/06 10:50	..
Boat Sampler	TOC Standard	2017/01/06 10:54	
Boat Sampler	Sample	2017/01/06 11:01	
Boat Sampler	...Sample	.. 2017/01/06 11:07	..
Boat Sampler	Sample	2017/01/06 11:10	
Boat Sampler	Sample	2017/01/06 11:13	
Boat Sampler	...Sample	.. 2017/01/06 11:19	..
Boat Sampler	Sample	2017/01/06 11:22	
Boat Sampler	Sample	2017/01/06 11:25	
Boat Sampler	...Sample	.. 2017/01/06 11:29	..
Boat Sampler	Sample	2017/01/06 11:36	
Boat Sampler	Sample	2017/01/06 11:39	
Boat Sampler	...Sample	.. 2017/01/06 11:42	..
Boat Sampler	Sample	2017/01/06 11:44	
Boat Sampler	Sample	2017/01/06 11:50	
Boat Sampler	...Sample	.. 2017/01/06 11:53	..
Boat Sampler	Sample	2017/01/06 11:57	
Boat Sampler	Sample	2017/01/06 12:00	
Boat Sampler	...Sample	.. 2017/01/06 12:03	..
Boat Sampler	Sample	2017/01/06 12:06	
Boat Sampler	Sample	2017/01/06 12:08	
Boat Sampler	...Sample	.. 2017/01/06 12:11	..
Boat Sampler	Sample	2017/01/06 12:15	
Boat Sampler	Sample	2017/01/06 12:17	

PA

1/6/17

Cal. Curve ID: TOC SOIL
Created: 2017/01/06 12:51
Calibration Factor (m): 7.515e+04
Y Intercept (b): 85899
r-squared: 0.99930

Standard ID	Y	X Expected	Measured	Message	Date & Time
BLANK	Raw Data	ug C	ug C		
	51902	0.000	-0.452		2017/01/06 10:24
250mg/l	836527	10.000	9.988		2017/01/06 10:28
500mg/l	1562927	20.000	19.654		2017/01/06 10:33
1000mg/l	3200398	40.000	41.444		2017/01/06 10:50
2000mg/l	6050278	80.000	79.366		2017/01/06 10:54

CHEMTECH

SOP ID: M9060/A-TOC

QA Control Code: A2040083

BALANCE ID :

WC SC-1

Sample ID	Weight (mg)	Result (PPM)	Average-Result (PPM)
LB85213BLS	40	8	10
LB85213BLS	40	14	
LB85213BLS	40	10	
LB85213BLS	40	9	
LB85213BSS	40	956	964
LB85213BSS	40	1013	
LB85213BSS	40	970	
LB85213BSS	40	915	
H6103-01	40	52	54
H6103-01	40	59	
H6103-01	40	51	
H6103-01	40	53	
H6103-02	40	553	572
H6103-02	40	585	
H6103-02	40	546	
H6103-02	40	602	