



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85216

Temp.: 16.7 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2033
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/06/2017	13:30
2	CAL2	1	Water			7.01	01/06/2017	13:31
3	CAL3	1	Water			10.04	01/06/2017	13:32
4	ICV	1	Water			7.02	01/06/2017	13:33
5	CCV1	1	Water			2.02	01/06/2017	13:34
6	I1060-01	1	Water			4.92	01/06/2017	13:35
7	I1060-01DUP	1	Water			4.93	01/06/2017	13:36
8	CCV2	1	Water			12.01	01/06/2017	13:37

2B85216

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i 1060

WorkList ID : 94183

Date : 1/6/2017 1:17:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1060-01	pH	Cool 4 deg C	STER06	D11	POOLWATER	01/04/2017	SM 4500-PH B

Date/Time 01-06-17 1:17 PM
 Received by: 711
 Relinquished by: [Signature]

Date/Time 01-06-17 2:30 PM
 Received by: [Signature]
 Relinquished by: JD