

1B85218

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/6/2017 11:34

Test: Phenolics-

Sample Id	Result <i>ng/L</i>	Dil. 1 +	Response	Errors
ICV1	0.941	0.0	0.109	
ICB1	0.005	0.0	0.005	
CCV1	0.939	0.0	0.109	
CCB1	-0.000	0.0	0.005	
PB95772BLS	0.001	0.0	0.005	
I1025-03	0.002	0.0	0.005	
I1025-03D	0.003	0.0	0.005	
I1025-03S	0.992	0.0	0.115	
I1025-03SD	0.999	0.0	0.116	
PB95772BSS	0.935	0.0	0.109	
CCV2	0.952	0.0	0.111	
CCB2	-0.006	0.0	0.004	
N	12			
Mean	0.480			
SD	0.5012			
CV%	104.36			

Aquakem v. 7.2AQ1

Results from time period:

Fri Jan 06 10:40:24 2017

Fri Jan 06 11:32:22 2017

AK

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0512	mg/l	1/6/2017 10:40:24	
0.1PPM	A	Phenolics-	P	0.1033	mg/l	1/6/2017 10:40:25	
0.25PPM	A	Phenolics-	P	0.2454	mg/l	1/6/2017 10:40:26	
0.50PPM	A	Phenolics-	P	0.5012	mg/l	1/6/2017 10:40:27	
1.0PPM	A	Phenolics-	P	0.9976	mg/l	1/6/2017 10:40:28	
2.0PPM	A	Phenolics-	P	2.0013	mg/l	1/6/2017 10:40:29	
ICV1	S	Phenolics-	P	0.9413	mg/l	1/6/2017 11:05:23	
ICB1	S	Phenolics-	P	0.0046	mg/l	1/6/2017 11:05:24	
CCV1	S	Phenolics-	P	0.9392	mg/l	1/6/2017 11:05:25	
CCB1	S	Phenolics-	P	-0.0002	mg/l	1/6/2017 11:05:26	
PB95772BLS	S	Phenolics-	P	0.0011	mg/l	1/6/2017 11:05:27	
I1025-03	S	Phenolics-	P	0.0017	mg/l	1/6/2017 11:05:29	
I1025-03D	S	Phenolics-	P	0.0031	mg/l	1/6/2017 11:05:30	
I1025-03S	S	Phenolics-	P	0.9917	mg/l	1/6/2017 11:05:31	
I1025-03SD	S	Phenolics-	P	0.9993	mg/l	1/6/2017 11:05:32	
PB95772BSS	S	Phenolics-	P	0.9351	mg/l	1/6/2017 11:32:20	
CCV2	S	Phenolics-	P	0.9524	mg/l	1/6/2017 11:32:21	
CCB2	S	Phenolics-	P	-0.0057	mg/l	1/6/2017 11:32:22	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/6/2017 10:42


Test Phenolics-

Accepted 1/6/2017 10:42

Factor 8.986
Bias 0.005

Coeff. of det. 0.999985

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0512	0.0500	
2	Phenol-2	0.016	0.1033	0.1000	
3	Phenol-2	0.032	0.2454	0.2500	
4	Phenol-2	0.061	0.5012	0.5000	
5	Phenol-2	0.116	0.9976	1.0000	
6	Phenol-2	0.227	2.0013	2.0000	