

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17

Analyst : AK

Data File : LB85218.csv

Approved By : 

Approved Date : 1/9/18

Worksheet # :

m-9065 - Phenolics - 11

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Phenol												
0.05PPM	0.05PPM				W	1/6/17						
Phenolics	PASS		0.051	0.051							mg/L	
0.1PPM	0.1PPM				W	1/6/17						
Phenolics	PASS		0.103	0.103							mg/L	
0.25PPM	0.25PPM				W	1/6/17						
Phenolics	PASS		0.245	0.245							mg/L	
0.50PPM	0.50PPM				W	1/6/17						
Phenolics	PASS		0.501	0.501							mg/L	
1.0PPM	1.0PPM				W	1/6/17						
Phenolics	PASS		0.998	0.998							mg/L	
2.0PPM	2.0PPM				W	1/6/17						
Phenolics	PASS		2.001	2.00							mg/L	
ICV1	ICV1				W	1/6/17						
Phenolics	PASS		0.941	0.94		94.0	90	110			mg/L	
ICB1	ICB1				W	1/6/17						
Phenolics	PASS		0.005	0.005				+/-0.0100			mg/L	
CCV1	CCV1				W	1/6/17						
Phenolics	PASS		0.939	0.94		94.0	90	110			mg/L	
CCB1	CCB1				W	1/6/17						
Phenolics	PASS		0.000	0.000				+/-0.0100			mg/L	
PB95772BLS	PB95772BLS				S	1/6/17						
Phenolics	PASS		0.001	0.055				+/-0.0780			mg/Kg	
I1025-03	TP7-8-COMP10		1		S	1/6/17						
Phenolics	PASS		0.002	0.130							mg/Kg	
I1025-03D	TP7-8-COMP10D		1		S	1/6/17						
Phenolics	FAIL		0.003	0.239					59	20	mg/Kg	
I1025-03S	TP7-8-COMP10S		1		S	1/6/17						
Phenolics	PASS		0.992	76.500		99.0	75	125			mg/Kg	
I1025-03SD	TP7-8-COMP10SD		1		S	1/6/17						
Phenolics	PASS		0.999	77.800		100.0	75	125	2	20	mg/Kg	
PB95772BSS	PB95772BSS				S	1/6/17						
Phenolics	PASS		0.935	46.80		94.0	80.00	120.00			mg/Kg	
CCV2	CCV2				W	1/6/17						
Phenolics	PASS		0.952	0.95		95.0	90	110			mg/L	
CCB2	CCB2				W	1/6/17						
Phenolics	PASS		-0.006	-0.006				+/-0.0100			mg/L	