

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AK
Data File : LB85219.csv

Approved By : 
Approved Date : 1/9/17
Worksheet # :

msm 4500 N Org B or C TKN-7

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Line 2
TKN											
0.5PPM	0.5PPM			W	1/6/17						
TKN		PASS	0.594	0.594							mg/L
1.0PPM	1.0PPM			W	1/6/17						
TKN		PASS	1.057	1.06							mg/L
2.5PPM	2.5PPM			W	1/6/17						
TKN		PASS	2.479	2.48							mg/L
5.0PPM	5.0PPM			W	1/6/17						
TKN		PASS	4.878	4.88							mg/L
6.7PPM	6.7PPM			W	1/6/17						
TKN		PASS	6.475	6.47							mg/L
10.0PPM	10.0PPM			W	1/6/17						
TKN		PASS	10.184	10.2							mg/L
ICV1	ICV1			W	1/6/17						
TKN		PASS	4.945	4.95	99.0	80	120				mg/L
ICB1	ICB1			W	1/6/17						
TKN		PASS	0.162	0.162					+/-0.0960		mg/L
CCV1	CCV1			W	1/6/17						
TKN		PASS	5.134	5.13	103.0	80	120				mg/L
CCB1	CCB1			W	1/6/17						
TKN		PASS	0.160	0.160					+/-0.0960		mg/L
PB95769BLW	PB95769BLW			W	1/6/17						
TKN		PASS	0.164	0.164					+/-0.0960		mg/L
PB95769BSW	PB95769BSW			W	1/6/17						
TKN		PASS	5.102	5.10	102.0	80	120				mg/L
I1003-01	DSN001		1	W	1/6/17						
TKN		PASS	0.404	0.404							mg/L
I1003-01D	DSN001D		1	W	1/6/17						
TKN		PASS	0.407	0.410						1.5	20 mg/L
I1003-01S	DSN001S		1	W	1/6/17						
TKN		PASS	5.591	5.590	104.0	75	125				mg/L
I1003-01SD	DSN001SD		1	W	1/6/17						
TKN		PASS	5.557	5.560	103.0	75	125	0.5	20		mg/L
CCV2	CCV2			W	1/6/17						
TKN		PASS	4.964	4.96	99.0	80	120				mg/L
CCB2	CCB2			W	1/6/17						
TKN		PASS	0.158	0.158					+/-0.0960		mg/L