

WinHg Database 1.7

FileUtilityHelp

RN↓

RN↑

?

INSTRUMENT ID: CV2

Protocolchemtech

Dataset/ProtoLB85220 /chemtech

ProtocolLine infoCal CurveReportCtrl ChartViewer

Reset

Calib Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B2.09514e-4

C6.39469e-3

Rho.999702

TypeLinear

Print Cal

☒ Calibrated

☒ Accepted

☐ Accepted

☐ New

μ Abs.48168

06Jan-17 14:41

Conc.10.1

IncludeS6Rep 1☒2☐3☐4☐5☐

S	Conc.	Calc.	Dev.	Mean	SD or %R...	Rep 1	Rep 2	Rep 3	Rep 4	F
01	.00000	-.007	-.007	-63	0	-63				
02	.20000	.191	-.009	880	0%	879				
03	2.5000	2.56	.065	12212	0%	12212				
04	5.0000	5.04	.036	24006	0%	24005				
05	7.5000	7.32	-.183	34894	0%	34894				
06	10.000	10.1	.098	48168	0%	48168				

Ready

CAPNUM

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 0		14:30:25	06 Jan 17	HG
Hg	.000	PPB	-63					
*** Standard: 2 Rep: 1				Seq: 1		14:32:26	06 Jan 17	HG
Hg	.200	PPB	879					
*** Standard: 3 Rep: 1				Seq: 2		14:35:00	06 Jan 17	HG
Hg	2.50	PPB	12212					
*** Standard: 4 Rep: 1				Seq: 3		14:37:03	06 Jan 17	HG
Hg	5.00	PPB	24005					
*** Standard: 5 Rep: 1				Seq: 4		14:39:08	06 Jan 17	HG
Hg	7.50	PPB	34894					
*** Standard: 6 Rep: 1				Seq: 5		14:41:19	06 Jan 17	HG
Hg	10.0	PPB	48168					
*** Sample ID: ICB115				Seq: 6		14:44:24	06 Jan 17	HG
Hg	3.98	PPB	18944					
*** Sample ID: ICB115				Seq: 7		14:46:27	06 Jan 17	HG
Hg	-.020	PPB	-124					
*** Sample ID: CCV133				Seq: 8		14:48:29	06 Jan 17	HG
Hg	5.01	PPB	23865					
*** Sample ID: CCB133				Seq: 9		14:50:30	06 Jan 17	HG
Hg	-.018	PPB	-115					
*** Sample ID: CRA				Seq: 10		14:52:30	06 Jan 17	HG
Hg	.200	PPB	923					

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Sample ID: HIGH STD				Seq: 11	14:54:43	06 Jan 17	HG	
			HIGH STD					
Hg	10.0	PPB	47910					=
=====								
*** Sample ID: CHK STD				Seq: 12	14:56:53	06 Jan 17	HG	
			CHK STD					
Hg	6.98	PPB	33273					=
=====								
*** Sample ID: PBW				Seq: 13	14:58:55	06 Jan 17	HG	
			PB95811BL					
Hg	-.035	PPB	-199					=
=====								
*** Sample ID: LCSW				Seq: 14	15:00:59	06 Jan 17	HG	
			PB95811BS					
Hg	3.60	PPB	17141					=
=====								
*** Sample ID: I1002-08				Seq: 15	15:03:33	06 Jan 17	HG	
			MDL-WATER-1					
Hg	.116	PPB	523					=
=====								
*** Sample ID: I1002-09				Seq: 16	15:05:42	06 Jan 17	HG	
			MDL-WATER-2					
Hg	.107	PPB	481					=
=====								
*** Sample ID: I1002-10				Seq: 17	15:07:52	06 Jan 17	HG	
			MDL-WATER-3					
Hg	.104	PPB	464					=
=====								
*** Sample ID: I1002-11				Seq: 18	15:10:25	06 Jan 17	HG	
			MDL-WATER-4					
Hg	.100	PPB	449					=
=====								
*** Sample ID: I1002-12				Seq: 19	15:12:23	06 Jan 17	HG	
			MDL-WATER-5					
Hg	.098	PPB	437					=
=====								
*** Sample ID: CCV134				Seq: 20	15:14:32	06 Jan 17	HG	
			CCV134					
Hg	5.04	PPB	24022					=
=====								

Protocol: chemtech

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Sample ID: CCB134				Seq: 21	15:16:51	06 Jan 17	HG	
			CCB134					
Hg	-.012	PPB	-89					
=====								
*** Sample ID: I1002-13				Seq: 22	15:18:50	06 Jan 17	HG	
			MDL-WATER-6					
Hg	.095	PPB	422					
=====								
*** Sample ID: I1002-14				Seq: 23	15:20:59	06 Jan 17	HG	
			MDL-WATER-7					
Hg	.127	PPB	574					
=====								
*** Sample ID: CCV135				Seq: 24	15:23:00	06 Jan 17	HG	
			CCV135					
Hg	5.27	PPB	25104					
=====								
*** Sample ID: CCB135				Seq: 25	15:25:19	06 Jan 17	HG	
			CCB135					
Hg	-.001	PPB	-34					
=====								