



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB85222

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP51923
calibration std. phosphate 0.5 ppm	WP51922
calibration std. phosphate 0.3 ppm	WP51921
calibration std. 0 ppm	WP51918
phosphate ICV std.	WP51924
phosphate CCV std.	WP51925
calibration std. phosphate 0.1 ppm	WP51920
calibration std. phosphate 0.05 ppm	WP51919
Combined reagent	WP51927
11 N sulfuric acid	WP48785
Phenolphthalein indicator	WP48972
Sodium hydroxide, 1N	WP48973

Intercept: -0

Slope: 0.6539

Regression: 0.999991

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/06/2017	12:30
2	CAL2	0.05	1	50	50	0.032	01/06/2017	12:30
3	CAL3	0.10	1	50	50	0.065	01/06/2017	12:31
4	CAL4	0.30	1	50	50	0.196	01/06/2017	12:31
5	CAL5	0.50	1	50	50	0.329	01/06/2017	12:32
6	CAL6	1.00	1	50	50	0.653	01/06/2017	12:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB85222

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.316	0.483	01/06/2017	12:33
2	ICB		1	50	50	0.000	0.000	01/06/2017	12:33
3	CCV1	0.50	1	50	50	0.326	0.499	01/06/2017	12:34
4	CCB1		1	50	50	0.000	0.000	01/06/2017	12:34
5	PB95803BL		1	50	50	0.000	0.000	01/06/2017	12:35
6	PB95803BS	0.50	1	50	50	0.312	0.477	01/06/2017	12:35
7	I1003-01		1	50	50	0.028	0.043	01/06/2017	12:36
8	I1003-01DUP		1	50	50	0.029	0.044	01/06/2017	12:36
9	I1003-01MS	0.50	1	50	50	0.350	0.535	01/06/2017	12:37
10	I1003-01MSD	0.50	1	50	50	0.348	0.532	01/06/2017	12:37
11	I1037-01		1	50	50	1.296	1.982	01/06/2017	12:38
12	I1037-01		10	50	50	0.306	0.468	01/06/2017	12:38
13	CCV2	0.50	1	50	50	0.327	0.500	01/06/2017	12:39
14	CCB2		1	50	50	0.000	0.000	01/06/2017	12:39