



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85227

Temp.: 22.4 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2033
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/06/2017	15:01
2	CAL2	1	Water			7.01	01/06/2017	15:04
3	CAL3	1	Water			10.02	01/06/2017	15:05
4	ICV	1	Water			7.02	01/06/2017	15:07
5	CCV1	1	Water			2.02	01/06/2017	15:08
6	I1057-08	1	Solid	20.02	20	10.91	01/06/2017	15:09
7	I1057-09	1	Solid	20.01	20	10.89	01/06/2017	15:10
8	I1057-10	1	Solid	20.04	20	10.90	01/06/2017	15:11
9	I1057-11	1	Solid	20.02	20	10.91	01/06/2017	15:12
10	I1057-12	1	Solid	20.04	20	10.96	01/06/2017	15:13
11	I1057-13	1	Solid	20.03	20	10.93	01/06/2017	15:15
12	I1057-14	1	Solid	20.01	20	10.95	01/06/2017	15:16
13	I1064-11	1	Solid	20.00	20	9.12	01/06/2017	15:17
14	I1064-13	1	Solid	20.03	20	8.57	01/06/2017	15:18
15	I1073-02	1	Solid	20.04	20	5.19	01/06/2017	15:20
16	CCV2	1	Water			12.01	01/06/2017	15:21
17	I1073-02DUP	1	Solid	20.02	20	5.20	01/06/2017	15:22
18	CCV3	1	Water			12.02	01/06/2017	15:24

WORKLIST(Hardcopy Internal Chain)

1985227

WorkList Name : CORROSIONITY I 1057

WorkList ID : 94179

Date : 1/6/2017 12:50:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1057-08	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP3	01/06/2017	9045C
	Solid	I1057-09	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP4	01/06/2017	9045C
	Solid	I1057-10	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP5	01/06/2017	9045C
	Solid	I1057-11	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP6	01/06/2017	9045C
	Solid	I1057-12	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP7	01/06/2017	9045C
	Solid	I1057-13	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP8	01/06/2017	9045C
	Solid	I1057-14	Corrosivity	Cool 4 deg C	SKAN06	D41	P3-SP9	01/06/2017	9045C
	Solid	I1064-11	Corrosivity	Cool 4 deg C	DEWB01	D41	TP5-6-COMP11	01/06/2017	9045C
	Solid	I1064-13	Corrosivity	Cool 4 deg C	DEWB01	D41	TP4-COMP12	01/06/2017	9045C
01/11/2017	Solid	I1073-02	Corrosivity	Cool 4 deg C	PSEG01	D31	L-SWITCH-SP-0-2-COMP	01/06/2017	9045C

Date/Time 01-06-17 12:50 PM
 Received by: JP
 Relinquished by: IR

Date/Time 01-06-17 5:30 PM
 Received by: SR
 Relinquished by: JP