

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85229.csv

Approved By : 
 Approved Date : 1/10/17
 Worksheet # :

M300.0 - Inorganic Arsenic - 16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Line 2
Bromide											
STD1	STD1			W	12/19/16						
Bromide		PASS		0.000	0.000						mg/L
STD2	STD2			W	12/19/16						
Bromide		PASS		0.982	0.982						mg/L
STD3	STD3			W	12/19/16						
Bromide		PASS		2.326	2.33						mg/L
STD4	STD4			W	12/19/16						
Bromide		PASS		4.181	4.18						mg/L
STD5	STD5			W	12/19/16						
Bromide		PASS		9.734	9.73						mg/L
STD6	STD6			W	12/19/16						
Bromide		PASS		19.435	19.4						mg/L
STD7	STD7			W	12/19/16						
Bromide		PASS		24.257	24.3						mg/L
STD8	STD8			W	12/19/16						
Bromide		PASS		50.585	50.6						mg/L
ICV1	ICV1			W	12/19/16						
Bromide		PASS		24.524	24.50	98.0		90	110		mg/L
ICB1	ICB1			W	12/19/16						
Bromide		PASS		0.000	0.000				+/-0.0660		mg/L
CCV1	CCV1			W	1/6/17						
Bromide		PASS		25.902	25.90	104.0		90	110		mg/L
CCB1	CCB1			W	1/6/17						
Bromide		PASS		0.000	0.000				+/-0.0660		mg/L
LB85229BLW	LB85229BLW			W	1/6/17						
Bromide		PASS		0.000	0.000				+/-0.0660		mg/L
LB85229BSW	LB85229BSW			W	1/6/17						
Bromide		PASS		25.833	25.80	103.0		90	110		mg/L
I1060-01	POOLWATER		1	W	1/6/17						
Bromide		PASS		0.000	0.000						mg/L
I1060-01D	POOLWATERD		1	W	1/6/17						
Bromide		PASS		0.000	0.00					0	20
I1060-01S	POOLWATERS		1	W	1/6/17						
Bromide		PASS		24.487	24.50	98.0		80	120		mg/L
CCV2	CCV2			W	1/6/17						
Bromide		PASS		25.317	25.30	101.0		90	110		mg/L
CCB2	CCB2			W	1/6/17						
Bromide		PASS		0.000	0.000				+/-0.0660		mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85229.csv


Approved By : 
 Approved Date : 1/10/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Chloride											
STD1	STD1				W	12/19/16					
Chloride		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/19/16					
Chloride		PASS	0.383		0.383						mg/L
STD3	STD3				W	12/19/16					
Chloride		PASS	0.770		0.770						mg/L
STD4	STD4				W	12/19/16					
Chloride		PASS	1.266		1.27						mg/L
STD5	STD5				W	12/19/16					
Chloride		PASS	2.897		2.90						mg/L
STD6	STD6				W	12/19/16					
Chloride		PASS	5.700		5.70						mg/L
STD7	STD7				W	12/19/16					
Chloride		PASS	7.167		7.17						mg/L
STD8	STD8				W	12/19/16					
Chloride		PASS	15.268		15.3						mg/L
ICV1	ICV1				W	12/19/16					
Chloride		PASS	7.282		7.28	97.0	90	110			mg/L
ICB1	ICB1				W	12/19/16					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
CCV1	CCV1				W	1/6/17					
Chloride		PASS	7.500		7.50	100.0	90	110			mg/L
CCB1	CCB1				W	1/6/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85229BLW	LB85229BLW				W	1/6/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85229BSW	LB85229BSW				W	1/6/17					
Chloride		PASS	7.551		7.55	101.0	90	110			mg/L
I1060-01	POOLWATER			1	W	1/6/17					
Chloride		PASS	0.294		0.294						mg/L
I1060-01D	POOLWATERD			1	W	1/6/17					
Chloride		PASS	0.305		0.31			6.67	20		mg/L
I1060-01S	POOLWATERS			1	W	1/6/17					
Chloride		PASS	7.176		7.18	92.0	80	120			mg/L
CCV2	CCV2				W	1/6/17					
Chloride		PASS	7.438		7.44	99.0	90	110			mg/L
CCB2	CCB2				W	1/6/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
 Analyst : AP
 Data File : LB85229.csv

Approved By : 
 Approved Date : 1/12/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Fluoride										
STD1	STD1				W	12/19/16				
Fluoride		PASS	0.000		0.000					mg/L
STD2	STD2				W	12/19/16				
Fluoride		PASS	0.308		0.308					mg/L
STD3	STD3				W	12/19/16				
Fluoride		PASS	0.491		0.491					mg/L
STD4	STD4				W	12/19/16				
Fluoride		PASS	0.847		0.847					mg/L
STD5	STD5				W	12/19/16				
Fluoride		PASS	1.940		1.94					mg/L
STD6	STD6				W	12/19/16				
Fluoride		PASS	3.885		3.88					mg/L
STD7	STD7				W	12/19/16				
Fluoride		PASS	4.615		4.62					mg/L
STD8	STD8				W	12/19/16				
Fluoride		PASS	10.215		10.2					mg/L
ICV1	ICV1				W	12/19/16				
Fluoride		PASS	4.649		4.65	93.0	90	110		mg/L
ICB1	ICB1				W	12/19/16				
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L
CCV1	CCV1				W	1/6/17				
Fluoride		PASS	5.132		5.13	103.0	90	110		mg/L
CCB1	CCB1				W	1/6/17				
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L
LB85229BLW	LB85229BLW				W	1/6/17				
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L
LB85229BSW	LB85229BSW				W	1/6/17				
Fluoride		PASS	5.022		5.02	100.0	90	110		mg/L
I1060-01	POOLWATER			1	W	1/6/17				
Fluoride		PASS	0.000		0.000					mg/L
I1060-01D	POOLWATERD			1	W	1/6/17				
Fluoride		PASS	0.000		0.00			0	20	mg/L
I1060-01S	POOLWATERS			1	W	1/6/17				
Fluoride		PASS	4.839		4.84	97.0	80	120		mg/L
CCV2	CCV2				W	1/6/17				
Fluoride		PASS	5.120		5.12	102.0	90	110		mg/L
CCB2	CCB2				W	1/6/17				
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L