


7471A

LB 85237
INSTRUMENT ID: CV4

Linear

 μ Abs.:

44813



A= 0.0000e+000

B= 2.2396e-004

C= 1.0805e-002

Rho= 0.9998651

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	7	0.000	7				
0.2	0.200	0.218	0.018	926	0.0 %	926				
2.5	2.500	2.548	0.048	11329	0.0 %	11329				
5.0	5.000	4.870	-0.130	21695	0.0 %	21695				
7.5	7.500	7.505	0.005	33463	0.0 %	33463				
10.0	10.000	10.047	0.047	44813	0.0 %	44813				

LB85237 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 09 Jan 2017 11:45:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7471A	PPB	09 Jan 2017 11:47:56	S	Std
0.2 - 1	50.2	926	-	0.2000	7471A	PPB	09 Jan 2017 11:50:14	S	Std
2.5 - 1	52.5	11329	-	2.5000	7471A	PPB	09 Jan 2017 11:52:32	S	Std
5.0 - 1	55	21695	-	5.0000	7471A	PPB	09 Jan 2017 11:54:49	S	Std
7.5 - 1	57.5	33463	-	7.5000	7471A	PPB	09 Jan 2017 11:57:06	S	Std
10.0 - 1	510	44813	-	10.0000	7471A	PPB	09 Jan 2017 11:59:22	S	Std
ICV89 - 1	ICV89	16846	3.7836	-	7471A	PPB	09 Jan 2017 12:02:04	U	SMPL
ICB89 - 1	ICB89	-37	0.0025	-	7471A	PPB	09 Jan 2017 12:04:20	U	SMPL
CCV29 - 1	CCV29	22262	4.9965	-	7471A	PPB	09 Jan 2017 12:06:36	U	SMPL
CCB29 - 1	CCB29	-37	0.0025	-	7471A	PPB	09 Jan 2017 12:08:52	U	SMPL
CRA - 1	CRA	937	0.2207	-	7471A	PPB	09 Jan 2017 12:11:07	U	SMPL
HighStd - 1	HighStd	45001	10.0890	-	7471A	PPB	09 Jan 2017 12:13:23	U	SMPL
ChkStd - 1	ChkStd	30567	6.8564	-	7471A	PPB	09 Jan 2017 12:15:39	U	SMPL
PB95802BL - 1	PBS	-70	-0.0049	-	7471A	PPB	09 Jan 2017 12:17:55	U	SMPL
PB95802BS - 1	LCSS	14884	3.3442	-	7471A	PPB	09 Jan 2017 12:20:12	U	SMPL
I1025-03 - 1	TP7-8-COMP10	66524	14.9092	-	7471A	PPB	09 Jan 2017 12:22:29	U	SMPL
I1045-01 - 1	P3-SP1	4972	1.1243	-	7471A	PPB	09 Jan 2017 12:24:47	U	SMPL
I1045-02 - 1	P3-SP2	12453	2.7997	-	7471A	PPB	09 Jan 2017 12:27:05	U	SMPL
I1045-03 - 1	RAP-SP1	1493	0.3452	-	7471A	PPB	09 Jan 2017 12:29:23	U	SMPL
I1045-04 - 1	RAP-SP2	3009	0.6847	-	7471A	PPB	09 Jan 2017 12:31:39	U	SMPL
CCV30 - 1	CCV30	21840	4.9020	-	7471A	PPB	09 Jan 2017 12:33:55	U	SMPL
CCB30 - 1	CCB30	-48	0.0001	-	7471A	PPB	09 Jan 2017 12:36:12	U	SMPL
I1045-05 - 1	RAP-SP3	7565	1.7050	-	7471A	PPB	09 Jan 2017 12:38:28	U	SMPL
I1045-06 - 1	RAP-SP4	5702	1.2878	-	7471A	PPB	09 Jan 2017 12:40:44	U	SMPL
I1050-01 - 1	POSTEX-28-20170105	12622	2.8376	-	7471A	PPB	09 Jan 2017 12:43:00	U	SMPL
I1050-01DUP - 1	POSTEX-28-20170105DUP	10883	2.4481	-	7471A	PPB	09 Jan 2017 12:45:16	U	SMPL
I1050-01MS - 1	POSTEX-28-20170105MS	26081	5.8518	-	7471A	PPB	09 Jan 2017 12:47:33	U	SMPL
I1050-01MSD - 1	POSTEX-28-20170105MSD	26432	5.9304	-	7471A	PPB	09 Jan 2017 12:49:50	U	SMPL
I1050-02 - 1	POSTEX-29-20170105	39969	8.9621	-	7471A	PPB	09 Jan 2017 12:52:07	U	SMPL
I1050-01LX5 - 1		2286	0.5228	-	7471A	PPB	09 Jan 2017 12:54:24	U	SMPL
I1050-01A - 1		29974	6.7236	-	7471A	PPB	09 Jan 2017 12:56:42	U	SMPL
I1025-03DLX5 - 1		13176	2.9616	-	7471A	PPB	09 Jan 2017 12:58:58	U	SMPL
CCV31 - 1	CCV31	22166	4.9750	-	7471A	PPB	09 Jan 2017 13:01:15	U	SMPL
CCB31 - 1	CCB31	-56	-0.0017	-	7471A	PPB	09 Jan 2017 13:03:31	U	SMPL