

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17

Analyst : AK *AK*

Data File : LB85241.csv

Approved By : *[Signature]*

Approved Date : 1/10/17

Worksheet # :

M 300.0 - Inorganics Ammonia - 16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	RPD	Max RPD	Units	Line 2
Nitrite											
STD1	STD1			W	12/19/16						
Nitrite		PASS	0.000	0.000						mg/L	
STD2	STD2			W	12/19/16						
Nitrite		PASS	0.386	0.386						mg/L	
STD3	STD3			W	12/19/16						
Nitrite		PASS	0.762	0.762						mg/L	
STD4	STD4			W	12/19/16						
Nitrite		PASS	1.274	1.27						mg/L	
STD5	STD5			W	12/19/16						
Nitrite		PASS	2.901	2.90						mg/L	
STD6	STD6			W	12/19/16						
Nitrite		PASS	5.805	5.80						mg/L	
STD7	STD7			W	12/19/16						
Nitrite		PASS	7.329	7.33						mg/L	
STD8	STD8			W	12/19/16						
Nitrite		PASS	15.455	15.5						mg/L	
ICV1	ICV1			W	12/19/16						
Nitrite		PASS	7.352	7.35	98.0	90	110			mg/L	
ICB1	ICB1			W	12/19/16						
Nitrite		PASS	0.000	0.000			+/-0.0220			mg/L	
CCV1	CCV1			W	1/4/17						
Nitrite		PASS	7.470	7.47	98.0	90	110			mg/L	
CCB1	CCB1			W	1/4/17						
Nitrite		PASS	0.000	0.000			+/-0.0220			mg/L	
LB85241BLW	LB85241BLW			W	1/4/17						
Nitrite		PASS	0.000	0.000			+/-0.0220			mg/L	
LB85241BSW	LB85241BSW			W	1/4/17						
Nitrite		PASS	7.601	7.60	101.0	90	110			mg/L	
I1002-08	MDL-WATER-1		1	W	1/4/17						
Nitrite		PASS	0.189	0.189						mg/L	
I1002-09	MDL-WATER-2		1	W	1/4/17						
Nitrite		PASS	0.200	0.200						mg/L	
I1002-10	MDL-WATER-3		1	W	1/4/17						
Nitrite		PASS	0.192	0.192						mg/L	
I1002-11	MDL-WATER-4		1	W	1/4/17						
Nitrite		PASS	0.204	0.204						mg/L	
I1002-12	MDL-WATER-5		1	W	1/4/17						
Nitrite		PASS	0.186	0.186						mg/L	
I1002-13	MDL-WATER-6		1	W	1/4/17						
Nitrite		PASS	0.190	0.190						mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
Analyst : AK
Data File : LB85241.csv

Approved By : 
Approved Date : 1/10/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method				Line 2
Nitrite											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Nitrite	PASS	0.196		0.196						mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Nitrite	PASS	0.198		0.198						mg/L	
CCV2	CCV2			W	1/4/17						
Nitrite	PASS	7.733		7.73	102.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Nitrite	PASS	0.000		0.000			+/-0.0220			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
 Analyst : AK *AK*
 Data File : LB85241.csv

Approved By : *[Signature]*
 Approved Date : *1/10/17*
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
										Line 2
Ortho Phosphate										
STD1	STD1			W	12/19/16					
Ortho-Phosphate		PASS		0.000	0.000					mg/L
STD2	STD2			W	12/19/16					
Ortho-Phosphate		PASS		0.285	0.285					mg/L
STD3	STD3			W	12/19/16					
Ortho-Phosphate		PASS		1.157	1.16					mg/L
STD4	STD4			W	12/19/16					
Ortho-Phosphate		PASS		2.057	2.06					mg/L
STD5	STD5			W	12/19/16					
Ortho-Phosphate		PASS		4.821	4.82					mg/L
STD6	STD6			W	12/19/16					
Ortho-Phosphate		PASS		9.910	9.91					mg/L
STD7	STD7			W	12/19/16					
Ortho-Phosphate		PASS		11.261	11.3					mg/L
STD8	STD8			W	12/19/16					
Ortho-Phosphate		PASS		24.488	24.5					mg/L
ICV1	ICV1			W	12/19/16					
Ortho-Phosphate		PASS		11.501	11.50	92.0	110			mg/L
ICB1	ICB1			W	12/19/16					
Ortho-Phosphate		PASS		0.000	0.000		+/-0.0100			mg/L
CCV1	CCV1			W	1/4/17					
Ortho-Phosphate		PASS		12.539	12.50	103.0	90	110		mg/L
CCB1	CCB1			W	1/4/17					
Ortho-Phosphate		PASS		0.000	0.000		+/-0.0100			mg/L
LB85241BLW	LB85241BLW			W	1/4/17					
Ortho-Phosphate		PASS		0.000	0.000		+/-0.0100			mg/L
LB85241BSW	LB85241BSW			W	1/4/17					
Ortho-Phosphate		PASS		12.720	12.70	102.0	90	110		mg/L
I1002-08	MDL-WATER-1		1	W	1/4/17					
Ortho-Phosphate		PASS		0.151	0.151					mg/L
I1002-09	MDL-WATER-2		1	W	1/4/17					
Ortho-Phosphate		PASS		0.131	0.131					mg/L
I1002-10	MDL-WATER-3		1	W	1/4/17					
Ortho-Phosphate		PASS		0.106	0.106					mg/L
I1002-11	MDL-WATER-4		1	W	1/4/17					
Ortho-Phosphate		PASS		0.127	0.127					mg/L
I1002-12	MDL-WATER-5		1	W	1/4/17					
Ortho-Phosphate		PASS		0.104	0.104					mg/L
I1002-13	MDL-WATER-6		1	W	1/4/17					
Ortho-Phosphate		PASS		0.112	0.112					mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17

Analyst : AK

Data File : LB85241.csv

Approved By : 

Approved Date : 1/10/17

Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB				Method	Method				Line 2
			Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	
Ortho Phosphate											
I1002-14	MDL-WATER-7		1	W	1/4/17						
Ortho-Phosphate	PASS	0.153	0.153							mg/L	
I1002-14RE	MDL-WATER-7RE		1	W	1/4/17						
Ortho-Phosphate	PASS	0.151	0.151							mg/L	
CCV2	CCV2			W	1/4/17						
Ortho-Phosphate	PASS	13.080	13.10		108.0	90	110			mg/L	
CCB2	CCB2			W	1/4/17						
Ortho-Phosphate	PASS	0.000	0.000				+/-0.0100			mg/L	