


7470A

LB84254  
INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

33909



A= 0.0000e+000

B= 2.9248e-004

C= 7.3426e-004

Rho= 0.9998659

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.001 | 0.001  | 1     | 0.000      | 1     |       |       |       |       |
| 0.2    | 0.200  | 0.195 | -0.005 | 665   | 0.0 %      | 665   |       |       |       |       |
| 2.5    | 2.500  | 2.473 | -0.027 | 8452  | 0.0 %      | 8452  |       |       |       |       |
| 5.0    | 5.000  | 4.992 | -0.008 | 17067 | 0.0 %      | 17067 |       |       |       |       |
| 7.5    | 7.500  | 7.620 | 0.120  | 26051 | 0.0 %      | 26051 |       |       |       |       |
| 10.0   | 10.000 | 9.918 | -0.082 | 33909 | 0.0 %      | 33909 |       |       |       |       |

## LB85254 INSTRUMENT ID : CV1

| Method: 7470A   |                      | Operator: Admin |         | Date of Analysis: 10 Jan 2017 11:44:13 |       |                      |      |      |
|-----------------|----------------------|-----------------|---------|----------------------------------------|-------|----------------------|------|------|
| Sample ID       | Extended ID          | $\mu$ Abs.      | Conc.   | Std ConcMethod                         | Units | Date                 | Type | Type |
| 0.0 - 1         | 50                   | 1               | -       | 0.00007470A                            | PPB   | 10 Jan 2017 11:48:56 | S    | Std  |
| 0.2 - 1         | 50.2                 | 665             | -       | 0.20007470A                            | PPB   | 10 Jan 2017 11:51:05 | S    | Std  |
| 2.5 - 1         | 52.5                 | 8452            | -       | 2.50007470A                            | PPB   | 10 Jan 2017 11:53:13 | S    | Std  |
| 5.0 - 1         | 55                   | 17067           | -       | 5.00007470A                            | PPB   | 10 Jan 2017 11:55:21 | S    | Std  |
| 7.5 - 1         | 57.5                 | 26051           | -       | 7.50007470A                            | PPB   | 10 Jan 2017 11:57:29 | S    | Std  |
| 10.0 - 1        | 510                  | 33909           | -       | 10.00007470A                           | PPB   | 10 Jan 2017 12:04:27 | S    | Std  |
| ICV90 - 1       | ICV90                | 14199           | 4.1536  | -7470A                                 | PPB   | 10 Jan 2017 12:06:55 | U    | SMPL |
| ICB90 - 1       | ICB90                | -42             | -0.0115 | -7470A                                 | PPB   | 10 Jan 2017 12:09:01 | U    | SMPL |
| CCV32 - 1       | CCV32                | 17132           | 5.0115  | -7470A                                 | PPB   | 10 Jan 2017 12:11:08 | U    | SMPL |
| CCB32 - 1       | CCB32                | -41             | -0.0113 | -7470A                                 | PPB   | 10 Jan 2017 12:13:15 | U    | SMPL |
| CRA - 1         | CRA                  | 634             | 0.1862  | -7470A                                 | PPB   | 10 Jan 2017 12:15:21 | U    | SMPL |
| HighStd - 1     | HighStd              | 34573           | 10.1126 | -7470A                                 | PPB   | 10 Jan 2017 12:17:28 | U    | SMPL |
| ChkStd - 1      | ChkStd               | 24747           | 7.2387  | -7470A                                 | PPB   | 10 Jan 2017 12:19:34 | U    | SMPL |
| PB95841BL - 1   | PBW                  | -54             | -0.0151 | -7470A                                 | PPB   | 10 Jan 2017 12:21:42 | U    | SMPL |
| PB95841BS - 1   | LCSW                 | 13393           | 3.9179  | -7470A                                 | PPB   | 10 Jan 2017 12:23:50 | U    | SMPL |
| PB95841TB - 1   | PB95822TB            | -36             | -0.0098 | -7470A                                 | PPB   | 10 Jan 2017 12:25:58 | U    | SMPL |
| I1057-08 - 1    | P3-SP3               | -6              | -0.0010 | -7470A                                 | PPB   | 10 Jan 2017 12:28:06 | U    | SMPL |
| I1057-08DUP - 1 | P3-SP3DUP            | 0               | 0.0007  | -7470A                                 | PPB   | 10 Jan 2017 12:30:14 | U    | SMPL |
| I1057-08MS - 1  | P3-SP3MS             | 13082           | 3.8269  | -7470A                                 | PPB   | 10 Jan 2017 12:32:23 | U    | SMPL |
| I1057-08MSD - 1 | P3-SP3MSD            | 12726           | 3.7228  | -7470A                                 | PPB   | 10 Jan 2017 12:34:30 | U    | SMPL |
| CCV33 - 1       | CCV33                | 16967           | 4.9632  | -7470A                                 | PPB   | 10 Jan 2017 12:36:37 | U    | SMPL |
| CCB33 - 1       | CCB33                | -38             | -0.0104 | -7470A                                 | PPB   | 10 Jan 2017 12:38:44 | U    | SMPL |
| I1057-09 - 1    | P3-SP4               | 3               | 0.0016  | -7470A                                 | PPB   | 10 Jan 2017 12:40:51 | U    | SMPL |
| I1057-10 - 1    | P3-SP5               | -3              | -0.0001 | -7470A                                 | PPB   | 10 Jan 2017 12:42:58 | U    | SMPL |
| I1057-11 - 1    | P3-SP6               | -14             | -0.0034 | -7470A                                 | PPB   | 10 Jan 2017 12:45:05 | U    | SMPL |
| I1057-12 - 1    | P3-SP7               | 8               | 0.0031  | -7470A                                 | PPB   | 10 Jan 2017 12:47:11 | U    | SMPL |
| I1057-13 - 1    | P3-SP8               | 10              | 0.0037  | -7470A                                 | PPB   | 10 Jan 2017 12:49:19 | U    | SMPL |
| I1057-14 - 1    | P3-SP9               | 4               | 0.0019  | -7470A                                 | PPB   | 10 Jan 2017 12:51:27 | U    | SMPL |
| I1064-11 - 1    | TP5-6-COMP11         | 209             | 0.0619  | -7470A                                 | PPB   | 10 Jan 2017 12:53:35 | U    | SMPL |
| I1064-13 - 1    | TP4-COMP12           | -8              | -0.0016 | -7470A                                 | PPB   | 10 Jan 2017 12:55:43 | U    | SMPL |
| I1073-02 - 1    | L-SWITCH-SP-0-2-COMP | 4               | 0.0019  | -7470A                                 | PPB   | 10 Jan 2017 12:57:52 | U    | SMPL |
| CCV34 - 1       | CCV34                | 17062           | 4.9910  | -7470A                                 | PPB   | 10 Jan 2017 12:59:59 | U    | SMPL |
| CCB34 - 1       | CCB34                | -33             | -0.0089 | -7470A                                 | PPB   | 10 Jan 2017 13:02:06 | U    | SMPL |
| I1057-08LX5 - 1 |                      | -14             | -0.0034 | -7470A                                 | PPB   | 10 Jan 2017 13:04:12 | U    | SMPL |
| I1057-08A - 1   |                      | 15031           | 4.3970  | -7470A                                 | PPB   | 10 Jan 2017 13:06:19 | U    | SMPL |
| CCV35 - 1       | CCV35                | 17248           | 5.0454  | -7470A                                 | PPB   | 10 Jan 2017 13:08:26 | U    | SMPL |
| CCB35 - 1       | CCB35                | -36             | -0.0098 | -7470A                                 | PPB   | 10 Jan 2017 13:10:33 | U    | SMPL |