

245.1


LB85259

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

35769



A= 0.0000e+000

B= 2.8045e-004

C= 3.1179e-002

Rho= 0.9999307

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	9	0.000	9				
0.2	0.200	0.213	0.013	650	0.0 %	650				
2.5	2.500	2.497	-0.003	8792	0.0 %	8792				
5.0	5.000	4.933	-0.067	17478	0.0 %	17478				
7.5	7.500	7.460	-0.040	26490	0.0 %	26490				
10.0	10.000	10.063	0.063	35769	0.0 %	35769				

LB85259 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Jan 2017 14:53:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	245.1	PPB	10 Jan 2017 14:56:05	S	Std
0.2 - 1	50.2	650	-	0.2000	245.1	PPB	10 Jan 2017 14:58:13	S	Std
2.5 - 1	52.5	8792	-	2.5000	245.1	PPB	10 Jan 2017 15:00:21	S	Std
5.0 - 1	55	17478	-	5.0000	245.1	PPB	10 Jan 2017 15:02:29	S	Std
7.5 - 1	57.5	26490	-	7.5000	245.1	PPB	10 Jan 2017 15:04:37	S	Std
10.0 - 1	510	35769	-	10.0000	245.1	PPB	10 Jan 2017 15:06:45	S	Std
ICV91 - 1	ICV91	13495	3.8159	-245.1		PPB	10 Jan 2017 15:09:13	U	SMPL
ICB91 - 1	ICB91	-12	0.0278	-245.1		PPB	10 Jan 2017 15:11:19	U	SMPL
CCV36 - 1	CCV36	17389	4.9080	-245.1		PPB	10 Jan 2017 15:13:26	U	SMPL
CCB36 - 1	CCB36	-21	0.0253	-245.1		PPB	10 Jan 2017 15:15:32	U	SMPL
CRA - 1	CRA	653	0.2143	-245.1		PPB	10 Jan 2017 15:17:38	U	SMPL
HighStd - 1	HighStd	35402	9.9597	-245.1		PPB	10 Jan 2017 15:19:45	U	SMPL
ChkStd - 1	ChkStd	23756	6.6936	-245.1		PPB	10 Jan 2017 15:21:52	U	SMPL
PB95842BL - 1	PBW	-44	0.0188	-245.1		PPB	10 Jan 2017 15:24:00	U	SMPL
PB95842BS - 1	LCSW	14101	3.9858	-245.1		PPB	10 Jan 2017 15:26:07	U	SMPL
I1028-01 - 1	COMP	212	0.0906	-245.1		PPB	10 Jan 2017 15:28:15	U	SMPL
I1028-01DUP - 1	COMPDUP	219	0.0926	-245.1		PPB	10 Jan 2017 15:30:23	U	SMPL
I1028-01MS - 1	COMPMS	11700	3.3125	-245.1		PPB	10 Jan 2017 15:36:36	U	SMPL
I1028-01MSD - 1	COMPMSD	12563	3.5545	-245.1		PPB	10 Jan 2017 15:38:44	U	SMPL
CCV37 - 1	CCV37	17295	4.8816	-245.1		PPB	10 Jan 2017 15:40:51	U	SMPL
CCB37 - 1	CCB37	-22	0.0250	-245.1		PPB	10 Jan 2017 15:42:57	U	SMPL
I1028-01LX5 - 1		49	0.0449	-245.1		PPB	10 Jan 2017 15:45:04	U	SMPL
I1028-01A - 1		12679	3.5870	-245.1		PPB	10 Jan 2017 15:47:11	U	SMPL
CCV38 - 1	CCV38	18066	5.0978	-245.1		PPB	10 Jan 2017 15:49:17	U	SMPL
CCB38 - 1	CCB38	-35	0.0214	-245.1		PPB	10 Jan 2017 15:51:24	U	SMPL