


7471A

LB 85266
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45901



A= 0.0000e+000

B= 2.1863e-004

C= 3.3606e-003

Rho= 0.9999503

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	64	0.000	64				
0.2	0.200	0.189	-0.011	847	0.0 %	847				
2.5	2.500	2.496	-0.004	11402	0.0 %	11402				
5.0	5.000	5.031	0.031	22996	0.0 %	22996				
7.5	7.500	7.428	-0.072	33959	0.0 %	33959				
10.0	10.000	10.039	0.039	45901	0.0 %	45901				

LB85266 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 10 Jan 2017 16:00:58

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	64	-	0.0000	7471A	PPB	10 Jan 2017 16:02:56	S	Std
0.2 - 1	50.2	847	-	0.2000	7471A	PPB	10 Jan 2017 16:05:14	S	Std
2.5 - 1	52.5	11402	-	2.5000	7471A	PPB	10 Jan 2017 16:07:32	S	Std
5.0 - 1	55	22996	-	5.0000	7471A	PPB	10 Jan 2017 16:09:49	S	Std
7.5 - 1	57.5	33959	-	7.5000	7471A	PPB	10 Jan 2017 16:12:06	S	Std
10.0 - 1	59	45901	-	10.0000	7471A	PPB	10 Jan 2017 16:14:23	S	Std
ICV92 - 1	ICV92	17919	3.9211	-	7471A	PPB	10 Jan 2017 16:17:04	U	SMPL
ICB92 - 1	ICB92	-51	-0.0078	-	7471A	PPB	10 Jan 2017 16:19:20	U	SMPL
CCV39 - 1	CCV39	22386	4.8977	-	7471A	PPB	10 Jan 2017 16:21:35	U	SMPL
CCB39 - 1	CCB39	-61	-0.0100	-	7471A	PPB	10 Jan 2017 16:23:51	U	SMPL
CRA - 1	CRA	804	0.1791	-	7471A	PPB	10 Jan 2017 16:26:07	U	SMPL
HighStd - 1	HighStd	45954	10.0505	-	7471A	PPB	10 Jan 2017 16:28:23	U	SMPL
ChkStd - 1	ChkStd	31856	6.9682	-	7471A	PPB	10 Jan 2017 16:30:39	U	SMPL
PB95847BL - 1	PBS	-98	-0.0181	-	7471A	PPB	10 Jan 2017 16:32:55	U	SMPL
I1057-01 - 1	P3-SP3	16183	3.5415	-	7471A	PPB	10 Jan 2017 16:37:29	U	SMPL
I1057-02 - 1	P3-SP4	21192	4.6366	-	7471A	PPB	10 Jan 2017 16:39:46	U	SMPL
I1057-03 - 1	P3-SP5	15194	3.3253	-	7471A	PPB	10 Jan 2017 16:42:04	U	SMPL
I1057-04 - 1	P3-SP6	17231	3.7706	-	7471A	PPB	10 Jan 2017 16:44:21	U	SMPL
I1057-05 - 1	P3-SP7	16587	3.6298	-	7471A	PPB	10 Jan 2017 16:46:37	U	SMPL
CCV40 - 1	CCV40	22154	4.8470	-	7471A	PPB	10 Jan 2017 16:48:54	U	SMPL
CCB40 - 1	CCB40	-53	-0.0082	-	7471A	PPB	10 Jan 2017 16:51:10	U	SMPL
I1057-06 - 1	P3-SP8	14675	3.2118	-	7471A	PPB	10 Jan 2017 16:53:26	U	SMPL
I1057-07 - 1	P3-SP9	15077	3.2997	-	7471A	PPB	10 Jan 2017 16:55:42	U	SMPL
I1064-03 - 1	TP5-6-COMP11	47475	10.3830	-	7471A	PPB	10 Jan 2017 16:57:58	U	SMPL
I1064-09 - 1	TP4-COMP12	24231	5.3011	-	7471A	PPB	10 Jan 2017 17:00:14	U	SMPL
I1073-01 - 1	L-SWITCH-SP-0-2-COMP	1317	0.2913	-	7471A	PPB	10 Jan 2017 17:02:31	U	SMPL
I1073-01DUP - 1	L-SWITCH-SP-0-2-COMPDUP	1342	0.2968	-	7471A	PPB	10 Jan 2017 17:04:48	U	SMPL
I1073-01MS - 1	L-SWITCH-SP-0-2-COMPMS	17584	3.8478	-	7471A	PPB	10 Jan 2017 17:07:05	U	SMPL
I1073-01MSD - 1	L-SWITCH-SP-0-2-COMPMSD	17004	3.7210	-	7471A	PPB	10 Jan 2017 17:09:23	U	SMPL
I1078-01 - 1	ROLL-OFF-0-2-COMP	340	0.0777	-	7471A	PPB	10 Jan 2017 17:11:41	U	SMPL
I1079-01 - 1	USDOT-617122-COMP-0-2	430	0.0974	-	7471A	PPB	10 Jan 2017 17:13:57	U	SMPL
CCV41 - 1	CCV41	22267	4.8717	-	7471A	PPB	10 Jan 2017 17:16:13	U	SMPL
CCB41 - 1	CCB41	-41	-0.0056	-	7471A	PPB	10 Jan 2017 17:18:29	U	SMPL
PB95847BS - 1	LCSS	16887	3.6954	-	7471A	PPB	10 Jan 2017 17:20:46	U	SMPL
I1073-01LX5 - 1		244	0.0567	-	7471A	PPB	10 Jan 2017 17:23:02	U	SMPL
I1073-01A - 1		18921	4.1401	-	7471A	PPB	10 Jan 2017 17:25:18	U	SMPL
I1064-03DLX5 - 1		9436	2.0664	-	7471A	PPB	10 Jan 2017 17:27:35	U	SMPL
CCV42 - 1	CCV42	22451	4.9119	-	7471A	PPB	10 Jan 2017 17:29:51	U	SMPL
CCB42 - 1	CCB42	-36	-0.0045	-	7471A	PPB	10 Jan 2017 17:32:08	U	SMPL