

7470A


LB 85271

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38017



A= 0.0000e+000

B= 2.5939e-004

C= 1.2889e-002

Rho= 0.9995829

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-44	0.000	-44				
0.2	0.200	0.201	0.001	725	0.0 %	725				
2.5	2.500	2.485	-0.015	9532	0.0 %	9532				
5.0	5.000	4.923	-0.077	18931	0.0 %	18931				
7.5	7.500	7.715	0.215	29693	0.0 %	29693				
10.0	10.000	9.874	-0.126	38017	0.0 %	38017				

LB85271 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jan 2017 10:02:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-44	-	0.00007470A	PPB		11 Jan 2017 10:11:26	S	Std
0.2 - 1	50.2	725	-	0.20007470A	PPB		11 Jan 2017 10:13:34	S	Std
2.5 - 1	52.5	9532	-	2.50007470A	PPB		11 Jan 2017 10:15:42	S	Std
5.0 - 1	55	18931	-	5.00007470A	PPB		11 Jan 2017 10:17:50	S	Std
7.5 - 1	57.5	29693	-	7.50007470A	PPB		11 Jan 2017 10:19:58	S	Std
10.0 - 1	510	38017	-	10.00007470A	PPB		11 Jan 2017 10:22:06	S	Std
ICV93 - 1	ICV93	15084	3.9255	-7470A	PPB		11 Jan 2017 10:24:51	U	SMPL
ICB93 - 1	ICB93	-45	0.0012	-7470A	PPB		11 Jan 2017 10:26:57	U	SMPL
CCV43 - 1	CCV43	19225	4.9996	-7470A	PPB		11 Jan 2017 10:29:03	U	SMPL
CCB43 - 1	CCB43	-31	0.0048	-7470A	PPB		11 Jan 2017 10:31:10	U	SMPL
CRA - 1	CRA	776	0.2142	-7470A	PPB		11 Jan 2017 10:33:16	U	SMPL
HighStd - 1	HighStd	38115	9.8994	-7470A	PPB		11 Jan 2017 10:35:23	U	SMPL
ChkStd - 1	ChkStd	26965	7.0073	-7470A	PPB		11 Jan 2017 10:37:30	U	SMPL
PB95885BL - 1	PBW	-45	0.0012	-7470A	PPB		11 Jan 2017 10:39:37	U	SMPL
PB95885BS - 1	LCSW	14635	3.8090	-7470A	PPB		11 Jan 2017 10:41:45	U	SMPL
PB95885TB - 1	PB95845TB	-8	0.0108	-7470A	PPB		11 Jan 2017 10:43:52	U	SMPL
I1078-02 - 1	ROLL-OFF-0-2-COMP	-10	0.0103	-7470A	PPB		11 Jan 2017 10:46:00	U	SMPL
I1078-02DUP - 1	ROLL-OFF-0-2-COMPDUP	0	0.0129	-7470A	PPB		11 Jan 2017 10:48:08	U	SMPL
I1078-02MS - 1	ROLL-OFF-0-2-COMPMS	13484	3.5105	-7470A	PPB		11 Jan 2017 10:50:17	U	SMPL
I1078-02MSD - 1	ROLL-OFF-0-2-COMPMSD	13527	3.5216	-7470A	PPB		11 Jan 2017 10:56:54	U	SMPL
CCV44 - 1	CCV44	18858	4.9044	-7470A	PPB		11 Jan 2017 10:59:00	U	SMPL
CCB44 - 1	CCB44	-34	0.0041	-7470A	PPB		11 Jan 2017 11:01:06	U	SMPL
I1079-02 - 1	USDOT-617122-COMP-0-2	0	0.0129	-7470A	PPB		11 Jan 2017 11:03:12	U	SMPL
I1078-02LX5 - 1		-9	0.0106	-7470A	PPB		11 Jan 2017 11:05:19	U	SMPL
I1078-02A - 1		15944	4.1486	-7470A	PPB		11 Jan 2017 11:07:26	U	SMPL
CCV45 - 1	CCV45	18920	4.9205	-7470A	PPB		11 Jan 2017 11:09:33	U	SMPL
CCB45 - 1	CCB45	-31	0.0048	-7470A	PPB		11 Jan 2017 11:11:41	U	SMPL