


ISM02.3

LB 85277
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4499E



A= 0.0000e+000

B= 2.2264e-004

C= 2.0945e-002

Rho= 0.9999332

Accept=Accepted

slope
y = 10160e-004 x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	12	0.000	12					10
0.20	0.200	0.220	0.020	892	0.0 %	892					-2
2.50	2.500	2.439	-0.061	10860	0.0 %	10860					1
5.00	5.000	5.037	0.037	22532	0.0 %	22532					-1
7.50	7.500	7.442	-0.058	33333	0.0 %	33333					0
10.00	10.000	10.039	0.039	44995	0.0 %	44995					

mB
01/10/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	12	-		0 ISM02.3	PPB	1/11/2017 13:54	Std
	0.2 S0.2	892	-		0.2 ISM02.3	PPB	1/11/2017 13:56	Std
	2.5 S2.5	10860	-		2.5 ISM02.3	PPB	1/11/2017 13:59	Std
	5 S5	22532	-		5 ISM02.3	PPB	1/11/2017 14:01	Std
	7.5 S7.5	33333	-		7.5 ISM02.3	PPB	1/11/2017 14:03	Std
	10 S10	44995	-		10 ISM02.3	PPB	1/11/2017 14:05	Std
ICV	ICV	16866	3.7759 -		ISM02.3	PPB	1/11/2017 14:08	SMPL
ICB	ICB	-15	0.0176 -		ISM02.3	PPB	1/11/2017 14:10	SMPL
CCV046	CCV046	22157	4.9539 -		ISM02.3	PPB	1/11/2017 14:12	SMPL
CCB046	CCB046	-48	0.0103 -		ISM02.3	PPB	1/11/2017 14:14	SMPL
PB95886BL	PBW003	2	0.0214 -		ISM02.3	PPB	1/11/2017 14:16	SMPL
I1076-01	MESWS6	7	0.0225 -		ISM02.3	PPB	1/11/2017 14:19	SMPL
I1076-03	MESWS7	16	0.0245 -		ISM02.3	PPB	1/11/2017 14:21	SMPL
I1076-05	MESWS8	12	0.0236 -		ISM02.3	PPB	1/11/2017 14:23	SMPL
I1076-06	MESWS8D	-5	0.0198 -		ISM02.3	PPB	1/11/2017 14:25	SMPL
I1076-07	MESWS8S	4495	1.0217 -		ISM02.3	PPB	1/11/2017 14:27	SMPL
I1076-11	MESWT1	0	0.0209 -		ISM02.3	PPB	1/11/2017 14:29	SMPL
CCV047	CCV047	22195	4.9624 -		ISM02.3	PPB	1/11/2017 14:31	SMPL
CCB047	CCB047	-21	0.0163 -		ISM02.3	PPB	1/11/2017 14:33	SMPL