

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/11/17

Analyst : AK

Data File : lb85287.csv

Approved By : 

Approved Date : 1/11/17

Worksheet # :

SM4500 - CN, CFE
Total Amenable & Reactive cyanide 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/11/17						
Cyanide		PASS	0.939		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/11/17						
Cyanide		PASS	6.095		0.006						mg/L	
10PPBCN	10PPBCN				W	1/11/17						
Cyanide		PASS	10.332		0.010						mg/L	
50PPBCN	50PPBCN				W	1/11/17						
Cyanide		PASS	53.363		0.053						mg/L	
100PPBCN	100PPBCN				W	1/11/17						
Cyanide		PASS	99.548		0.100						mg/L	
250PPBCN	250PPBCN				W	1/11/17						
Cyanide		PASS	239.974		0.240						mg/L	
500PPBCN	500PPBCN				W	1/11/17						
Cyanide		PASS	504.750		0.505						mg/L	
ICV1	ICV1				W	1/11/17						
Cyanide		PASS	95.686		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/11/17						
Cyanide		PASS	1.599		0.002			+/-0.0010			mg/L	
CCV1	CCV1				W	1/11/17						
Cyanide		PASS	233.649		0.23	92.0	90	110			mg/L	
CCB1	CCB1				W	1/11/17						
Cyanide		PASS	1.552		0.002			+/-0.0010			mg/L	
PB95870BLW	PB95870BLW				W	1/11/17						
Cyanide		PASS	1.688		0.002			+/-0.0010			mg/L	
PB95870BSW	PB95870BSW				W	1/11/17						
Cyanide		PASS	95.633		0.096	96.0	85	115			mg/L	
MIDPB95870	MIDPB95870				W	1/11/17						
Cyanide		PASS	233.967		0.234						mg/L	
I1092-03	Annual -LOD- Water-2017			1	W	1/11/17						
Cyanide		PASS	1.315		0.001						mg/L	
I1092-04	Annual -LOQ- Water-2017				W	1/11/17						
Cyanide		PASS	8.992		0.009	90.0	50	150			mg/L	
CCV2	CCV2				W	1/11/17						
Cyanide		PASS	244.866		0.25	100.0	90	110			mg/L	
CCB2	CCB2				W	1/11/17						
Cyanide		PASS	1.259		0.001			+/-0.0010			mg/L	