


7471A

LB 85289
INSTRUMENT ID: CY1

Linear

 μ Abs.:

43910



A= 0.0000e+000

B= 2.2384e-004

C= -2.0848e-002

Rho= 0.9994272

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	5	0.000	5				
0.2	0.200	0.184	-0.016	915	0.0 %	915				
2.5	2.500	2.427	-0.073	10937	0.0 %	10937				
5.0	5.000	5.097	0.097	22863	0.0 %	22863				
7.5	7.500	7.704	0.204	34508	0.0 %	34508				
10.0	10.000	9.808	-0.192	43910	0.0 %	43910				

LB85289 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Jan 2017 14:49:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	5	-	0.0000	7471A	PPB	11 Jan 2017 14:51:24	S	Std
0.2 - 1	S0.2	915	-	0.2000	7471A	PPB	11 Jan 2017 14:53:42	S	Std
2.5 - 1	S2.5	10937	-	2.5000	7471A	PPB	11 Jan 2017 14:56:00	S	Std
5.0 - 1	S5	22863	-	5.0000	7471A	PPB	11 Jan 2017 14:58:18	S	Std
7.5 - 1	S7.5	34508	-	7.5000	7471A	PPB	11 Jan 2017 15:00:35	S	Std
10.0 - 1	S10	43910	-	10.0000	7471A	PPB	11 Jan 2017 15:05:35	S	Std
ICV94 - 1	ICV94	17621	3.9235	-	7471A	PPB	11 Jan 2017 15:08:28	U	SMPL
ICB94 - 1	ICB94	-19	-0.0251	-	7471A	PPB	11 Jan 2017 15:10:43	U	SMPL
CCV48 - 1	CCV48	22778	5.0778	-	7471A	PPB	11 Jan 2017 15:12:59	U	SMPL
CCB48 - 1	CCB48	-42	-0.0302	-	7471A	PPB	11 Jan 2017 15:15:15	U	SMPL
CRA - 1	CRA	916	0.1842	-	7471A	PPB	11 Jan 2017 15:17:31	U	SMPL
HighStd - 1	HighStd	44800	10.0073	-	7471A	PPB	11 Jan 2017 15:19:47	U	SMPL
ChkStd - 1	ChkStd	31342	6.9948	-	7471A	PPB	11 Jan 2017 15:22:03	U	SMPL
PB95905BL - 1	PBS	-47	-0.0314	-	7471A	PPB	11 Jan 2017 15:24:19	U	SMPL
PB95905BS - 1	LCSS	16226	3.6112	-	7471A	PPB	11 Jan 2017 15:26:36	U	SMPL
I1098-03 - 1	TP3-COMP13	39652	8.8550	-	7471A	PPB	11 Jan 2017 15:28:53	U	SMPL
I1098-03DUP - 1	TP3-COMP13DUP	42520	9.4969	-	7471A	PPB	11 Jan 2017 15:31:11	U	SMPL
I1098-03MS - 1	TP3-COMP13MS	54937	12.2764	-	7471A	PPB	11 Jan 2017 15:33:28	U	SMPL
I1098-03MSD - 1	TP3-COMP13MSD	55104	12.3138	-	7471A	PPB	11 Jan 2017 15:35:46	U	SMPL
CCV49 - 1	CCV49	22644	5.0478	-	7471A	PPB	11 Jan 2017 15:38:02	U	SMPL
CCB49 - 1	CCB49	-48	-0.0316	-	7471A	PPB	11 Jan 2017 15:40:18	U	SMPL
I1098-03LX5 - 1		7719	1.7070	-	7471A	PPB	11 Jan 2017 15:42:34	U	SMPL
I1098-03A - 1		56511	12.6287	-	7471A	PPB	11 Jan 2017 15:44:50	U	SMPL
I1098-03DLX5 - 1		7508	1.6598	-	7471A	PPB	11 Jan 2017 15:47:07	U	SMPL
I1098-03DUPDLX5 - 1		8339	1.8458	-	7471A	PPB	11 Jan 2017 15:49:23	U	SMPL
I1098-03MSDLX5 - 1		11029	2.4479	-	7471A	PPB	11 Jan 2017 15:51:39	U	SMPL
I1098-03MSDDLX5 - 1		11057	2.4542	-	7471A	PPB	11 Jan 2017 15:53:55	U	SMPL
I1098-03LDLX25 - 1		1545	0.3250	-	7471A	PPB	11 Jan 2017 15:56:12	U	SMPL
I1098-03ADLX5 - 1		11009	2.4434	-	7471A	PPB	11 Jan 2017 15:58:30	U	SMPL
CCV50 - 1	CCV50	22622	5.0429	-	7471A	PPB	11 Jan 2017 16:00:47	U	SMPL
CCB50 - 1	CCB50	-33	-0.0282	-	7471A	PPB	11 Jan 2017 16:03:05	U	SMPL

msb
1/13/17