



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85292

Temp.: 19.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/11/2017	15:10
2	CAL2	1	Water			7.01	01/11/2017	15:11
3	CAL3	1	Water			10.03	01/11/2017	15:12
4	ICV	1	Water			7.02	01/11/2017	15:13
5	CCV1	1	Water			2.02	01/11/2017	15:14
6	I1137-01	1	Water			4.67	01/11/2017	15:15
7	I1137-01DUP	1	Water			4.68	01/11/2017	15:16
8	CCV2	1	Water			12.01	01/11/2017	15:17

2B85292

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1137

WorkList ID : 94313

Date : 1/11/2017 2:50:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1137-01	pH	Cool 4 deg C	STER11	D21	POOLWATER	01/06/2017	SM 4500-PH B

Date/Time 01-11-17 2:50 PM
Received by: JP
Relinquished by: JP

Date/Time 01-11-17 3:55 PM
Received by: JP
Relinquished by: JP