


7470A

LB 85316  
INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

37924



A= 0.0000e+000

B= 2.6412e-004

C= 3.1809e-002

Rho= 0.9997274

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.032  | 0.032  | 2     | 0.000      | 2     |       |       |       |       |
| 0.2    | 0.200  | 0.217  | 0.017  | 703   | 0.0 %      | 703   |       |       |       |       |
| 2.5    | 2.500  | 2.554  | 0.054  | 9551  | 0.0 %      | 9551  |       |       |       |       |
| 5.0    | 5.000  | 4.809  | -0.191 | 18088 | 0.0 %      | 18088 |       |       |       |       |
| 7.5    | 7.500  | 7.539  | 0.039  | 28422 | 0.0 %      | 28422 |       |       |       |       |
| 10.0   | 10.000 | 10.048 | 0.048  | 37924 | 0.0 %      | 37924 |       |       |       |       |

## LB85316 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jan 2017 14:59:25

| Sample ID       | Extended ID   | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|---------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50            | 2      | -       | 0.0000   | 7470A  | PPB   | 12 Jan 2017 15:06:40 | S    | Std  |
| 0.2 - 1         | 50.2          | 703    | -       | 0.2000   | 7470A  | PPB   | 12 Jan 2017 15:08:48 | S    | Std  |
| 2.5 - 1         | 52.5          | 9551   | -       | 2.5000   | 7470A  | PPB   | 12 Jan 2017 15:10:56 | S    | Std  |
| 5.0 - 1         | 55            | 18088  | -       | 5.0000   | 7470A  | PPB   | 12 Jan 2017 15:13:05 | S    | Std  |
| 7.5 - 1         | 57.5          | 28422  | -       | 7.5000   | 7470A  | PPB   | 12 Jan 2017 15:15:13 | S    | Std  |
| 10.0 - 1        | 57.5          | 37924  | -       | 10.0000  | 7470A  | PPB   | 12 Jan 2017 15:19:52 | S    | Std  |
| ICV95 - 1       | ICV95         | 15171  | 4.0387  | -        | 7470A  | PPB   | 12 Jan 2017 15:22:21 | U    | SMPL |
| ICB95 - 1       | ICB95         | -37    | 0.0220  | -        | 7470A  | PPB   | 12 Jan 2017 15:24:27 | U    | SMPL |
| CCV51 - 1       | CCV51         | 18512  | 4.9211  | -        | 7470A  | PPB   | 12 Jan 2017 15:26:34 | U    | SMPL |
| CCB51 - 1       | CCB51         | -34    | 0.0228  | -        | 7470A  | PPB   | 12 Jan 2017 15:28:40 | U    | SMPL |
| CRA - 1         | CRA           | 712    | 0.2199  | -        | 7470A  | PPB   | 12 Jan 2017 15:30:47 | U    | SMPL |
| HighStd - 1     | HighStd       | 38339  | 10.1578 | -        | 7470A  | PPB   | 12 Jan 2017 15:32:53 | U    | SMPL |
| ChkStd - 1      | ChkStd        | 26703  | 7.0845  | -        | 7470A  | PPB   | 12 Jan 2017 15:35:00 | U    | SMPL |
| PB95934BL - 1   | PBW           | 4      | 0.0329  | -        | 7470A  | PPB   | 12 Jan 2017 15:37:07 | U    | SMPL |
| PB95934BS - 1   | LCSW          | 14350  | 3.8219  | -        | 7470A  | PPB   | 12 Jan 2017 15:39:15 | U    | SMPL |
| PB95934TB - 1   | PB95887TB     | 21     | 0.0374  | -        | 7470A  | PPB   | 12 Jan 2017 15:41:23 | U    | SMPL |
| I1098-05 - 1    | TP3-COMP13    | 45     | 0.0437  | -        | 7470A  | PPB   | 12 Jan 2017 15:43:31 | U    | SMPL |
| I1098-05DUP - 1 | TP3-COMP13DUP | 47     | 0.0442  | -        | 7470A  | PPB   | 12 Jan 2017 15:45:39 | U    | SMPL |
| I1098-05MS - 1  | TP3-COMP13MS  | 13363  | 3.5612  | -        | 7470A  | PPB   | 12 Jan 2017 15:47:48 | U    | SMPL |
| I1098-05MSD - 1 | TP3-COMP13MSD | 13939  | 3.7133  | -        | 7470A  | PPB   | 12 Jan 2017 15:49:54 | U    | SMPL |
| CCV52 - 1       | CCV52         | 18271  | 4.8575  | -        | 7470A  | PPB   | 12 Jan 2017 15:52:01 | U    | SMPL |
| CCB52 - 1       | CCB52         | -32    | 0.0234  | -        | 7470A  | PPB   | 12 Jan 2017 15:54:08 | U    | SMPL |
| I1098-05LX5 - 1 |               | 44     | 0.0434  | -        | 7470A  | PPB   | 12 Jan 2017 15:56:15 | U    | SMPL |
| I1098-05A - 1   |               | 15203  | 4.0472  | -        | 7470A  | PPB   | 12 Jan 2017 15:58:21 | U    | SMPL |
| CCV53 - 1       | CCV53         | 18224  | 4.8451  | -        | 7470A  | PPB   | 12 Jan 2017 16:00:28 | U    | SMPL |
| CCB53 - 1       | CCB53         | -38    | 0.0218  | -        | 7470A  | PPB   | 12 Jan 2017 16:02:35 | U    | SMPL |