



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85317

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP52093
COD CCV std, 50ppm	WP52096
COD LOD std, 1PPM	WP52080
COD LOQ std, 10.0PPM	WP52079

Temp In (C) : <u>150</u>	Date In: <u>01/12/2017</u>	Time In: <u>09:30</u>
Temp Out (C) : <u>150</u>	Date Out: <u>01/12/2017</u>	Time Out: <u>11:31</u>

Intercept: 0.2469

Slope: 0.9992

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	12/30/2016	12:50
2	CAL2	10	1	Water	10.000	12/30/2016	12:50
3	CAL3	50	1	Water	51.000	12/30/2016	12:51
4	CAL4	100	1	Water	100.000	12/30/2016	12:51
5	CAL5	150	1	Water	150.000	12/30/2016	12:52



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85317

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.793	01/12/2017	12:52
2	ICB				1	Water	0.000	-0.247	01/12/2017	12:53
3	CCV1	50			1	Water	50.000	49.793	01/12/2017	12:32
4	CCB1				1	Water	0.000	-0.247	01/12/2017	12:32
5	LB85317BL		1.00	50	1	Solid	0.000	-0.247	01/12/2017	12:33
6	LB85317BS	50	1.00	50	1	Solid	51.000	50.794	01/12/2017	12:33
7	I1092-01		1.01	50	1	Solid	1.000	0.754	01/12/2017	12:34
8	I1092-02		1.01	50	1	Solid	11.000	10.762	01/12/2017	12:34
9	CCV2	50			1	Water	51.000	50.794	01/12/2017	12:35
10	CCB2				1	Water	0.000	-0.247	01/12/2017	12:35

WORKLIST(Hardcopy Internal Chain)

21385317

Worklist Name : COD01/10/17

Worklist ID : 94254

Date : 1/10/2017 10:57:37 AM

Date/Time _____

Received by: _____

Relinquished by: _____

Date/Time _____

Received by: _____

Relinquished by: _____