



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85318

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/12/2017	15:25
2	CAL2	1	Water			7.01	01/12/2017	15:26
3	CAL3	1	Water			10.04	01/12/2017	15:27
4	ICV	1	Water			7.01	01/12/2017	15:30
5	CCV1	1	Water			2.02	01/12/2017	15:31
6	I1162-17	1	Solid	20.02	20	7.53	01/12/2017	15:32
7	I1162-17DUP	1	Solid	20.01	20	7.54	01/12/2017	15:33
8	I1162-18	1	Solid	20.04	20	7.85	01/12/2017	15:34
9	I1162-19	1	Solid	20.03	20	6.32	01/12/2017	15:35
10	I1162-20	1	Solid	20.00	20	7.27	01/12/2017	15:36
11	I1162-21	1	Solid	20.04	20	6.51	01/12/2017	15:37
12	I1162-22	1	Solid	20.05	20	8.20	01/12/2017	15:38
13	I1162-23	1	Solid	20.04	20	7.28	01/12/2017	15:40
14	I1162-24	1	Solid	20.03	20	7.30	01/12/2017	15:41
15	CCV2	1	Water			12.02	01/12/2017	15:42

LB 85318

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1163 WorkList ID : 94355 Date : 1/12/2017 3:07:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1162-17	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-03-COMP	01/10/2017	9045C
	Solid	I1162-18	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-45-COMP	01/09/2017	9045C
	Solid	I1162-19	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-46-COMP	01/09/2017	9045C
	Solid	I1162-20	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-47-COMP	01/09/2017	9045C
	Solid	I1162-21	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-48-COMP	01/09/2017	9045C
	Solid	I1162-22	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-49-COMP	01/09/2017	9045C
	Solid	I1162-23	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-08-COMP	01/11/2017	9045C
	Solid	I1162-24	Corrosivity	Cool 4 deg C	LIRO01	D22	SB-09-COMP	01/11/2017	9045C

Date/Time 01-12-17 3:00 PM
 Received by: J1'
 Relinquished by: CF

Date/Time 01-12-17 5:20 PM
 Received by: CP
 Relinquished by: J8