

Run Number: LB85326

DATE & TIME IN: 01/13/2017 10:00

OVEN TEMP: 104 °C

DATE & TIME OUT: 01/13/2017 14:15

OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB85326BL	LB85326BL	1.3852	1.3852	100	1.3852	1.3852	1.3852	0	0
2	LB85326BS	LB85326BS	1.4561	1.4561	100	1.5105	1.5105	1.5105	0.0544	544
3	I1144-02	FINAL-EFFLUENT-19C	1.4118	1.4118	100	1.4226	1.4226	1.4226	0.0108	108
4	I1157-01	SPDES-1	1.4071	1.4071	400	1.408	1.408	1.408	0.0009	2.3
5	I1157-01DUP	SPDES-1DUP	1.4104	1.4104	400	1.4111	1.4111	1.4111	0.0007	1.8
6	I1180-01	32220008-3	1.4109	1.4109	150	1.4264	1.4264	1.4264	0.0155	103.3
7	I1181-01	15220032-1	1.4084	1.4084	150	1.4418	1.4418	1.4418	0.0334	222.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB 85326

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I1157

WorkList ID : 94373

Date : 1/13/2017 8:46:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1144-02	TSS	Cool 4 deg C	M&MM01	D11	FINAL-EFFLUENT-19C	01/10/2017	SM2540 D
	Water	I1157-01	TSS	Cool 4 deg C	SKAN06	D22	SPDES-1	01/12/2017	SM2540 D
	Water	I1180-01	TSS	Cool 4 deg C	PHAR02	D22	32220008-3	01/12/2017	SM2540 D
	Water	I1181-01	TSS	Cool 4 deg C	PHAR02	D22	15220032-1	01/12/2017	SM2540 D

Date/Time 01-13-17 8:46 AM
 Received by:
 Relinquished by:

Date/Time 01-13-17 10:30 AM
 Received by:
 Relinquished by: