

AK 1/16/17

Sample Results Report Print Date/Time: 2017/01/13 12:53:30

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
CCV	10.3027			TOC	
CCB	0.3741			TOC	
LB85329BLW	0.3202	0.0070	2.19	TOC	
LB85329BSW.....	10.0254...	0.0322..	0.32...	TOC	..
I1092-03	0.2879	0.0503	17.49	TOC	
I1092-04	1.3155	0.0361	2.74	TOC	
CCV.....	9.8751...	..	...	TOC	..
CCB	0.3250			TOC	

AK 1/16/17

Sample Results Report Print Date/Time: 2017/01/13 12:53:30

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2017/01/13 09:38	
TOC 0 - 20 ppmC	Sample	3	2017/01/13 10:04	
TOC 0 - 20 ppmC	Sample	4	2017/01/13 10:20	
TOC 0 - 20 ppmC	...Sample	..	5..2017/01/13 10:56	..
TOC 0 - 20 ppmC	Sample	6	2017/01/13 11:31	
TOC 0 - 20 ppmC	Sample	7	2017/01/13 12:23	
TOC 0 - 20 ppmC	...Sample	..	2..2017/01/13 12:37	..
TOC 0 - 20 ppmC	Sample	3	2017/01/13 12:50	

Cal. Curve ID: TOC WATER 0-20PPM
Created: 2016/12/29 10:42
Calibration Factor (m): 6.890e+04
Y Intercept (b): 9954
r-squared: 0.99951

Standard ID	Y	X Expected	Measured	Message	Date & Time
	Raw Data	ug C	ug C		
0.0PPM	5652	0.000	-0.062		2016/12/29 09:07
0.5PPM	27564	0.250	0.256		2016/12/29 09:22
1.0PPM	45665	0.500	0.518		2016/12/29 09:35
2.0PPM	87909	1.000	1.131		2016/12/29 10:00
5.0PPM	180418	2.500	2.474		2016/12/29 10:13
10.0PPM	346324	5.000	4.882		2016/12/29 10:27
20.0PPM	702526	10.000	10.051		2016/12/29 10:41

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
0.0PPM	5652			TOC	
0.5PPM	27564			TOC	
1.0PPM	45665			TOC	
2.0PPM.....	87909...	..	...	TOC	..
5.0PPM	180418			TOC	
10.0PPM	346324			TOC	
20.0PPM.....	702526...	..	...	TOC	..
ICV	10.2906			TOC	
ICB	0.3396			TOC	

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2016/12/29 09:07	
TOC 0 - 20 ppmC	TOC Standard	2	2016/12/29 09:22	
TOC 0 - 20 ppmC	TOC Standard	3	2016/12/29 09:35	
TOC 0 - 20 ppmC	TOC Standard	4	2016/12/29 10:00	
TOC 0 - 20 ppmC	TOC Standard	5	2016/12/29 10:13	
TOC 0 - 20 ppmC	TOC Standard	6	2016/12/29 10:27	
TOC 0 - 20 ppmC	TOC Standard	7	2016/12/29 10:41	
TOC 0 - 20 ppmC	Sample	8	2016/12/29 10:55	
TOC 0 - 20 ppmC	Sample	9	2016/12/29 11:10	