

LB85329

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: SM5310B TOC
Parameter: TOC
Run Number: LB85329
Instrument: Appolo-9000

SM5310B - TOC - 10.

Analytical Summary Report

Analysis Method: SM5310B TOC
Parameter: TOC
Run Number: LB85329
Instrument: Appolo-9000

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	5652.000	1	12/29/16
2	0.5PPM	0.5PPM	CAL	27564.000	1	12/29/16
3	1.0PPM	1.0PPM	CAL	45665.000	1	12/29/16
4	2.0PPM	2.0PPM	CAL	87909.000	1	12/29/16
5	5.0PPM	5.0PPM	CAL	180418.000	1	12/29/16
6	10.0PPM	10.0PPM	CAL	346324.000	1	12/29/16
7	20.0PPM	20.0PPM	CAL	702526.000	1	12/29/16
8	ICV	ICV	ICV	10.291	1	12/29/16
9	ICB	ICB	ICB	0.000	1	12/29/16
10	CCV	CCV	CCV	10.303	1	1/13/17
11	CCB	CCB	CCB	0.000	1	1/13/17
12	LB85329BLW	LB85329BLW	MB	0.320	1	1/13/17
13	LB85329BSW	LB85329BSW	LCS	10.025	1	1/13/17
14	I1092-03	Annual -LOD- Water-2017	SAM	0.288	1	1/13/17
15	I1092-04	Annual -LOQ- Water-2017	SAM	1.316	1	1/13/17
16	CCV	CCV	CCV	9.875	1	1/13/17
17	CCB	CCB	CCB	0.000	1	1/13/17

* cal results are in row amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
Analyst : AK
Data File : LB85329.MDB

Approved By : 
Approved Date : 1/17/17
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
TOC										
0.0PPM	0.0PPM				W	12/29/16				
TOC		PASS	5652.000		5650					mg/L
0.5PPM	0.5PPM				W	12/29/16				
TOC		PASS	27564.000		27600					mg/L
1.0PPM	1.0PPM				W	12/29/16				
TOC		PASS	45665.000		45700					mg/L
2.0PPM	2.0PPM				W	12/29/16				
TOC		PASS	87909.000		87900					mg/L
5.0PPM	5.0PPM				W	12/29/16				
TOC		PASS	180418.000		180000					mg/L
10.0PPM	10.0PPM				W	12/29/16				
TOC		PASS	346324.000		346000					mg/L
20.0PPM	20.0PPM				W	12/29/16				
TOC		PASS	702526.000		703000					mg/L
ICV1	ICV1				W	12/29/16				
TOC		PASS	10.291		10.30	103.0	90	110		mg/L
ICB1	ICB1				W	12/29/16				
TOC		PASS	0.000		0.000			+/-0.0800		mg/L
CCV1	CCV1				W	1/13/17				
TOC		PASS	10.303		10.30	103.0	90	110		mg/L
CCB1	CCB1				W	1/13/17				
TOC		PASS	0.000		0.000			+/-0.0800		mg/L
LB85329BLW	LB85329BLW				W	1/13/17				
TOC		PASS	0.320		0.320			+/-0.0800		mg/L
LB85329BSW	LB85329BSW				W	1/13/17				
TOC		PASS	10.025		10.00	100.0	90	110		mg/L
I1092-03	Annual -LOD- Water-2017				W	1/13/17				
TOC		PASS	0.288		0.288					mg/L
I1092-04	Annual -LOQ- Water-2017				W	1/13/17				
TOC		PASS	1.316		1.32	132.0	50	150		mg/L
CCV2	CCV2				W	1/13/17				
TOC		PASS	9.875		9.88	99.0	90	110		mg/L
CCB2	CCB2				W	1/13/17				
TOC		PASS	0.000		0.000			+/-0.0800		mg/L

Cal results
are in
amount.