



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85333

Temp.: 18.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/13/2017	14:00
2	CAL2	1	Water			7.01	01/13/2017	14:01
3	CAL3	1	Water			10.02	01/13/2017	14:02
4	ICV	1	Water			7.03	01/13/2017	14:03
5	CCV1	1	Water			2.02	01/13/2017	14:05
6	I1194-01	1	Water			3.92	01/13/2017	14:06
7	I1194-01DUP	1	Water			3.93	01/13/2017	14:07
8	I1204-01	1	Water			3.63	01/13/2017	14:10
9	CCV2	1	Water			12.02	01/13/2017	14:11

LB85333

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1194

WorkList ID : 94385

Date : 1/13/2017 1:50:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1194-01	pH	Cool 4 deg C	STER12	D31	CELLPOOLWATER	01/12/2017	SM 4500-PH B
	Water	I1204-01	pH	Cool 4 deg C	STER04	D31	POOLWATER	01/06/2017	SM 4500-PH B

Date/Time 01-13-17 1:50 PM
Received by: JH
Relinquished by: J-L

Date/Time 01-13-17 3:30 PM
Received by: J-L
Relinquished by: JD