

ISM02.3


LB 85336

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47273



A= 0.0000e+000

B= 2.1210e-004

C= -4.8400e-003

Rho= 0.9999862

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	14	0.000	14					-1
0.20	0.200	0.198	-0.002	958	0.0 %	958					1
2.50	2.500	2.527	0.027	11937	0.0 %	11937					0
5.00	5.000	4.979	-0.021	23498	0.0 %	23498					0
7.50	7.500	7.475	-0.025	35267	0.0 %	35267					0
10.00	10.000	10.022	0.022	47273	0.0 %	47273					0

LB
1/13/17

LB85336
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	14	-		0 ISM02.3	PPB	1/13/2017 15:10	Std
	0.2 S0.2	958	-		0.2 ISM02.3	PPB	1/13/2017 15:12	Std
	2.5 S2.5	11937	-		2.5 ISM02.3	PPB	1/13/2017 15:14	Std
	5 S5	23498	-		5 ISM02.3	PPB	1/13/2017 15:16	Std
	7.5 S7.5	35267	-		7.5 ISM02.3	PPB	1/13/2017 15:18	Std
	10 S10	47273	-		10 ISM02.3	PPB	1/13/2017 15:20	Std
ICV	ICV	18546	3.9288 -		ISM02.3	PPB	1/13/2017 15:23	SMPL
ICB	ICB	-20	-0.0091 -		ISM02.3	PPB	1/13/2017 15:25	SMPL
CCV060	CCV060	23302	4.9376 -		ISM02.3	PPB	1/13/2017 15:27	SMPL
CCB060	CCB060	-19	-0.0089 -		ISM02.3	PPB	1/13/2017 15:29	SMPL
PB95962BL	PBS004	5	-0.0038 -		ISM02.3	PPB	1/13/2017 15:31	SMPL
I1129-01	ME5BJ5	2652	0.5577 -		ISM02.3	PPB	1/13/2017 15:34	SMPL
I1129-02	ME5BJ6	10990	2.3262 -		ISM02.3	PPB	1/13/2017 15:36	SMPL
I1129-03	ME5BJ7	16650	3.5267 -		ISM02.3	PPB	1/13/2017 15:38	SMPL
I1129-04	ME5BJ8	3945	0.8319 -		ISM02.3	PPB	1/13/2017 15:40	SMPL
I1129-05	ME5BJ9	2964	0.6238 -		ISM02.3	PPB	1/13/2017 15:42	SMPL
I1129-06	ME5BK0	3562	0.7507 -		ISM02.3	PPB	1/13/2017 15:44	SMPL
I1129-07	ME5BK1	2018	0.4232 -		ISM02.3	PPB	1/13/2017 15:46	SMPL
I1129-08	ME5BK2	4858	1.0256 -		ISM02.3	PPB	1/13/2017 15:48	SMPL
I1129-09	ME5BK3	4440	0.9369 -		ISM02.3	PPB	1/13/2017 15:51	SMPL
I1129-10	ME5BK4	4365	0.921 -		ISM02.3	PPB	1/13/2017 15:53	SMPL
I1129-11	ME5BK5	1664	0.3481 -		ISM02.3	PPB	1/13/2017 15:55	SMPL
I1129-12	ME5BK6	8173	1.7287 -		ISM02.3	PPB	1/13/2017 15:57	SMPL
I1129-13	ME5BK7	7306	1.5448 -		ISM02.3	PPB	1/13/2017 15:59	SMPL
I1129-14	ME5BK8	3627	0.7645 -		ISM02.3	PPB	1/13/2017 16:01	SMPL
I1129-15	ME5BK9	1385	0.2889 -		ISM02.3	PPB	1/13/2017 16:03	SMPL
I1129-16	ME5BL0	2579	0.5422 -		ISM02.3	PPB	1/13/2017 16:05	SMPL
I1129-17	ME5BL1	3444	0.7256 -		ISM02.3	PPB	1/13/2017 16:08	SMPL
I1129-18	ME5BL2	5850	1.236 -		ISM02.3	PPB	1/13/2017 16:10	SMPL
I1129-19	ME5BL3	15226	3.2246 -		ISM02.3	PPB	1/13/2017 16:12	SMPL
CCV061	CCV061	23603	5.0014 -		ISM02.3	PPB	1/13/2017 16:14	SMPL
CCB061	CCB061	-17	-0.0084 -		ISM02.3	PPB	1/13/2017 16:16	SMPL
I1129-20	ME5BL4	6571	1.3889 -		ISM02.3	PPB	1/13/2017 16:18	SMPL
I1129-21	ME5BL4D	6411	1.355 -		ISM02.3	PPB	1/13/2017 16:20	SMPL
I1129-22	ME5BL4S	18567	3.9333 -		ISM02.3	PPB	1/13/2017 16:22	SMPL
CCV062	CCV062	23331	4.9437 -		ISM02.3	PPB	1/13/2017 16:24	SMPL
CCB062	CCB062	-12	-0.0074 -		ISM02.3	PPB	1/13/2017 16:27	SMPL