



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85341

Temp.: 22.6 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/13/2017	18:30
2	CAL2	1	Water			7.02	01/13/2017	18:31
3	CAL3	1	Water			10.04	01/13/2017	18:32
4	ICV	1	Water			7.01	01/13/2017	18:33
5	CCV1	1	Water			2.02	01/13/2017	18:35
6	I1182-05	1	Solid	20.02	20	8.48	01/13/2017	18:36
7	I1182-05DUP	1	Solid	20.04	20	8.49	01/13/2017	18:37
8	I1219-01	1	Solid	20.01	20	7.39	01/13/2017	18:38
9	CCV2	1	Water			12.02	01/13/2017	18:40

WORKLIST(Hardcopy Internal Chain)

285341

WorkList Name : CORROSIONITY I1182

WorkList ID : 94386

Date : 1/13/2017 1:51:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/26/2017	Solid	I1182-05	Corrosivity	Cool 4 deg C	PARS15	D32	TOD-WCS-011217	01/12/2017	9045C
	Solid	I1219-01	Corrosivity	Cool 4 deg C	PARS15	D32	TOD-WCS-011317(11-47)	01/13/2017	9045C

Date/Time 01-13-17 11:24 AM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 01-13-17 7:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]