



Analytical Summary Report

Analysis Method: 9045C
Parameter: pH
Run Number: LB85342

Analyst By : JIGNESH
Supervisor Review By : apatel
Temp.: 21.2 °C
SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/13/2017	18:30
2	CAL2	1	Water			7.02	01/13/2017	18:31
3	CAL3	1	Water			10.04	01/13/2017	18:32
4	ICV	1	Water			7.01	01/13/2017	18:33
5	CCV1	1	Water			2.02	01/13/2017	18:34
6	I1219-03	1	Water			7.64	01/13/2017	18:45
7	I1219-03DUP	1	Water			7.65	01/13/2017	18:46
8	CCV2	1	Water			12.02	01/13/2017	18:47

WORKLIST(Hardcopy Internal Chain)

LB 85342

WorkList Name : PH I1219

WorkList ID : 94401

Date : 1/13/2017 6:26:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1219-03	pH	Cool 4 deg C	PARS15	D32	WC011217-LIQUID-TODT-HIL	01/12/2017	9045C

Date/Time 01-13-17 4:30 PM
 Received by: J11
 Relinquished by: _____

Date/Time 01-13-17 7:00 PM
 Received by: _____
 Relinquished by: J11