

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17

Analyst : AP

Data File : LB85348.csv

Approved By :

Approved Date :

Worksheet # :

M300-D - Inorganic Arsenic - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Bromide											
STD1	STD1				W	12/19/16					
Bromide		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/19/16					
Bromide		PASS	0.982		0.982						mg/L
STD3	STD3				W	12/19/16					
Bromide		PASS	2.326		2.33						mg/L
STD4	STD4				W	12/19/16					
Bromide		PASS	4.181		4.18						mg/L
STD5	STD5				W	12/19/16					
Bromide		PASS	9.734		9.73						mg/L
STD6	STD6				W	12/19/16					
Bromide		PASS	19.435		19.4						mg/L
STD7	STD7				W	12/19/16					
Bromide		PASS	24.257		24.3						mg/L
STD8	STD8				W	12/19/16					
Bromide		PASS	50.585		50.6						mg/L
ICV1	ICV1				W	12/19/16					
Bromide		PASS	24.524		24.50	98.0	90	110			mg/L
ICB1	ICB1				W	12/19/16					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
CCV1	CCV1				W	1/13/17					
Bromide		PASS	25.344		25.30	101.0	90	110			mg/L
CCB1	CCB1				W	1/13/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
LB85348BLW	LB85348BLW				W	1/13/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
LB85348BSW	LB85348BSW				W	1/13/17					
Bromide		PASS	25.076		25.10	100.0	90	110			mg/L
I1194-01	CELLPOOLWATER				W	1/13/17					
Bromide		PASS	0.000		0.000						mg/L
I1204-01	POOLWATER				W	1/13/17					
Bromide		PASS	0.000		0.000						mg/L
I1194-01D	CELLPOOLWATERD				W	1/13/17					
Bromide		PASS	0.000		0.00			0	20		mg/L
I1194-01S	CELLPOOLWATERS				W	1/13/17					
Bromide		PASS	24.862		24.90	100.0	80	120			mg/L
CCV2	CCV2				W	1/13/17					
Bromide		PASS	24.763		24.80	99.0	90	110			mg/L
CCB2	CCB2				W	1/13/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17

Analyst : AP

Data File : LB85348.csv

Approved By : 

Approved Date : 1/17/17


Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride												
STD1	STD1				W	12/19/16						
Chloride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	12/19/16						
Chloride		PASS	0.383		0.383						mg/L	
STD3	STD3				W	12/19/16						
Chloride		PASS	0.770		0.770						mg/L	
STD4	STD4				W	12/19/16						
Chloride		PASS	1.266		1.27						mg/L	
STD5	STD5				W	12/19/16						
Chloride		PASS	2.897		2.90						mg/L	
STD6	STD6				W	12/19/16						
Chloride		PASS	5.700		5.70						mg/L	
STD7	STD7				W	12/19/16						
Chloride		PASS	7.167		7.17						mg/L	
STD8	STD8				W	12/19/16						
Chloride		PASS	15.268		15.3						mg/L	
ICV1	ICV1				W	12/19/16						
Chloride		PASS	7.282		7.28	97.0	90	110			mg/L	
ICB1	ICB1				W	12/19/16						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
CCV1	CCV1				W	1/13/17						
Chloride		PASS	7.199		7.20	96.0	90	110			mg/L	
CCB1	CCB1				W	1/13/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB85348BLW	LB85348BLW				W	1/13/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB85348BSW	LB85348BSW				W	1/13/17						
Chloride		PASS	7.393		7.39	99.0	90	110			mg/L	
I1194-01	CELLPOOLWATER			1	W	1/13/17						
Chloride		PASS	0.000		0.000						mg/L	
I1204-01	POOLWATER			1	W	1/13/17						
Chloride		PASS	0.000		0.000						mg/L	
I1194-01D	CELLPOOLWATERD			1	W	1/13/17						
Chloride		PASS	0.000		0.00				0	20	mg/L	
I1194-01S	CELLPOOLWATERS			1	W	1/13/17						
Chloride		PASS	7.162		7.16	95.0	80	120			mg/L	
CCV2	CCV2				W	1/13/17						
Chloride		PASS	7.327		7.33	98.0	90	110			mg/L	
CCB2	CCB2				W	1/13/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AP
 Data File : LB85348.csv

Approved By : 
 Approved Date :
 Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Fluoride												
STD1	STD1				W	12/19/16						
Fluoride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	12/19/16						
Fluoride		PASS	0.308		0.308						mg/L	
STD3	STD3				W	12/19/16						
Fluoride		PASS	0.491		0.491						mg/L	
STD4	STD4				W	12/19/16						
Fluoride		PASS	0.847		0.847						mg/L	
STD5	STD5				W	12/19/16						
Fluoride		PASS	1.940		1.94						mg/L	
STD6	STD6				W	12/19/16						
Fluoride		PASS	3.885		3.88						mg/L	
STD7	STD7				W	12/19/16						
Fluoride		PASS	4.615		4.62						mg/L	
STD8	STD8				W	12/19/16						
Fluoride		PASS	10.215		10.2						mg/L	
ICV1	ICV1				W	12/19/16						
Fluoride		PASS	4.649		4.65	93.0	90	110			mg/L	
ICB1	ICB1				W	12/19/16						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
CCV1	CCV1				W	1/13/17						
Fluoride		PASS	5.255		5.26	105.0	90	110			mg/L	
CCB1	CCB1				W	1/13/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
LB85348BLW	LB85348BLW				W	1/13/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
LB85348BSW	LB85348BSW				W	1/13/17						
Fluoride		PASS	4.776		4.78	96.0	90	110			mg/L	
I1194-01	CELLPOOLWATER			1	W	1/13/17						
Fluoride		PASS	0.000		0.000						mg/L	
I1204-01	POOLWATER			1	W	1/13/17						
Fluoride		PASS	0.000		0.000						mg/L	
I1194-01D	CELLPOOLWATERD			1	W	1/13/17						
Fluoride		PASS	0.000		0.00				0	20	mg/L	
I1194-01S	CELLPOOLWATERS			1	W	1/13/17						
Fluoride		PASS	4.703		4.70	94.0	80	120			mg/L	
CCV2	CCV2				W	1/13/17						
Fluoride		PASS	4.825		4.82	96.0	90	110			mg/L	
CCB2	CCB2				W	1/13/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	