


245.1

LB85352
INSTRUMENT ID: CY1

Linear

 μ Abs.:

35082



A= 0.0000e+000

B= 2.8400e-004

C= 1.6466e-002

Rho= 0.9999742

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	-11	0.000	-11				
0.2	0.200	0.201	0.001	649	0.0 %	649				
2.5	2.500	2.484	-0.016	8687	0.0 %	8687				
5.0	5.000	4.972	-0.028	17449	0.0 %	17449				
7.5	7.500	7.551	0.051	26529	0.0 %	26529				
10.0	10.000	9.980	-0.020	35082	0.0 %	35082				

LB85352 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 16 Jan 2017 12:45:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	245.1	PPB	16 Jan 2017 12:55:04	S	Std
0.2 - 1	50.2	649	-	0.2000	245.1	PPB	16 Jan 2017 12:57:12	S	Std
2.5 - 1	52.5	8687	-	2.5000	245.1	PPB	16 Jan 2017 12:59:20	S	Std
5.0 - 1	55	17449	-	5.0000	245.1	PPB	16 Jan 2017 13:01:29	S	Std
7.5 - 1	57.5	26529	-	7.5000	245.1	PPB	16 Jan 2017 13:03:37	S	Std
10.0 - 1	510	35082	-	10.0000	245.1	PPB	16 Jan 2017 13:05:44	S	Std
ICV97 - 1	ICV97	14027	3.9651	-245.1		PPB	16 Jan 2017 13:08:09	U	SMPL
ICB97 - 1	ICB97	-31	0.0225	-245.1		PPB	16 Jan 2017 13:10:15	U	SMPL
CCV63 - 1	CCV63	17501	4.9867	-245.1		PPB	16 Jan 2017 13:13:08	U	SMPL
CCB63 - 1	CCB63	-26	0.0091	-245.1		PPB	16 Jan 2017 13:15:14	U	SMPL
CRA - 1	CRA	655	0.2025	-245.1		PPB	16 Jan 2017 13:17:21	U	SMPL
HighStd - 1	HighStd	34776	9.8928	-245.1		PPB	16 Jan 2017 13:19:27	U	SMPL
ChkStd - 1	ChkStd	24175	6.8821	-245.1		PPB	16 Jan 2017 13:21:34	U	SMPL
PB95982BL - 1	PBW	-45	0.0037	-245.1		PPB	16 Jan 2017 13:23:41	U	SMPL
PB95982BS - 1	LCSW	14515	4.1387	-245.1		PPB	16 Jan 2017 13:25:48	U	SMPL
I1180-01 - 1	32220008-3	-7	0.0145	-245.1		PPB	16 Jan 2017 13:27:56	U	SMPL
I1180-01MS - 1	32220008-3MS	13615	3.8831	-245.1		PPB	16 Jan 2017 13:32:12	U	SMPL
I1180-01MSD - 1	32220008-3MSD	14092	4.0186	-245.1		PPB	16 Jan 2017 13:34:21	U	SMPL
I1206-01 - 1	011217-DSDC-F-D-M	87	0.0412	-245.1		PPB	16 Jan 2017 13:36:28	U	SMPL
CCV64 - 1	CCV64	17018	4.8496	-245.1		PPB	16 Jan 2017 13:38:35	U	SMPL
CCB64 - 1	CCB64	-27	0.0088	-245.1		PPB	16 Jan 2017 13:40:41	U	SMPL
I1206-02 - 1	011217-DSDC-F-U-M	100	0.0449	-245.1		PPB	16 Jan 2017 13:42:48	U	SMPL
I1180-01LX5 - 1		7	0.0185	-245.1		PPB	16 Jan 2017 13:54:43	U	SMPL
I1180-01A - 1		14627	4.1705	-245.1		PPB	16 Jan 2017 13:56:49	U	SMPL
I1180-01DUP - 1	32220008-3DUP	4	0.0176	-245.1		PPB	16 Jan 2017 14:05:42	U	SMPL
CCV65 - 1	CCV65	17492	4.9842	-245.1		PPB	16 Jan 2017 14:07:49	U	SMPL
CCB65 - 1	CCB65	-33	0.0071	-245.1		PPB	16 Jan 2017 14:09:56	U	SMPL