

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB85354

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP52150
calibration std. phosphate 0.5 ppm	WP52149
calibration std. phosphate 0.3 ppm	WP52148
calibration std. 0 ppm	WP52145
phosphate ICV std.	WP52151
calibration std. phosphate 0.1 ppm	WP52147
calibration std. phosphate 0.05 ppm	WP52146
Combined reagent	WP52154
11 N sulfuric acid	WP48785
Phenolphthalein indicator	WP48972
Sodium hydroxide, 1N	WP48973
phosphate CCV std.	WP52152

Intercept: -0.0002Slope: 0.6538Regression: 0.999962

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/16/2017	11:04
2	CAL2	0.05	1	50	50	0.032	01/16/2017	11:04
3	CAL3	0.10	1	50	50	0.064	01/16/2017	11:05
4	CAL4	0.30	1	50	50	0.195	01/16/2017	11:05
5	CAL5	0.50	1	50	50	0.331	01/16/2017	11:06
6	CAL6	1.00	1	50	50	0.652	01/16/2017	11:06



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB85354

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.488	01/16/2017	11:07
2	ICB		1	50	50	0.000	0.000	01/16/2017	11:07
3	CCV1	0.50	1	50	50	0.330	0.505	01/16/2017	11:16
4	CCB1		1	50	50	0.000	0.000	01/16/2017	11:16
5	PB96028BL		1	1.00	50	0.000	0.000	01/16/2017	11:17
6	PB96028BS	0.5	1	1.01	50	0.322	0.493	01/16/2017	11:17
7	I1092-01		1	1.01	50	0.006	0.009	01/16/2017	11:18
8	I1092-02		1	1.00	50	0.068	0.104	01/16/2017	11:18
9	CCV2	0.50	1	50	50	0.328	0.502	01/16/2017	11:19
10	CCB2		1	50	50	0.000	0.000	01/16/2017	11:19