


7470A

LB85356
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36698



A= 0.0000e+000

B= 2.6906e-004

C= 4.2803e-003

Rho= 0.9997191

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-11	0.000	-11				
0.2	0.200	0.187	-0.013	678	0.0 %	678				
2.5	2.500	2.471	-0.029	9167	0.0 %	9167				
5.0	5.000	4.990	-0.010	18530	0.0 %	18530				
7.5	7.500	7.673	0.173	28502	0.0 %	28502				
10.0	10.000	9.878	-0.122	36698	0.0 %	36698				

LB85356 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jan 2017 14:26:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	7470A	PPB	16 Jan 2017 14:28:24	S	Std
0.2 - 1	50.2	678	-	0.2000	7470A	PPB	16 Jan 2017 14:30:32	S	Std
2.5 - 1	52.5	9167	-	2.5000	7470A	PPB	16 Jan 2017 14:32:41	S	Std
5.0 - 1	55	18530	-	5.0000	7470A	PPB	16 Jan 2017 14:34:49	S	Std
7.5 - 1	57.5	28502	-	7.5000	7470A	PPB	16 Jan 2017 14:36:56	S	Std
10.0 - 1	510	36698	-	10.0000	7470A	PPB	16 Jan 2017 14:39:03	S	Std
ICV98 - 1	ICV98	14625	3.9393	-7470A	PPB	PPB	16 Jan 2017 14:41:36	U	SMPL
ICB98 - 1	ICB98	-30	-0.0038	-7470A	PPB	PPB	16 Jan 2017 14:43:42	U	SMPL
CCV66 - 1	CCV66	18599	5.0085	-7470A	PPB	PPB	16 Jan 2017 14:45:48	U	SMPL
CCB66 - 1	CCB66	-20	-0.0011	-7470A	PPB	PPB	16 Jan 2017 14:47:55	U	SMPL
CRA - 1	CRA	757	0.2080	-7470A	PPB	PPB	16 Jan 2017 14:50:01	U	SMPL
HighStd - 1	HighStd	36859	9.9216	-7470A	PPB	PPB	16 Jan 2017 14:52:08	U	SMPL
ChkStd - 1	ChkStd	25859	6.9619	-7470A	PPB	PPB	16 Jan 2017 14:54:15	U	SMPL
PB95981BL - 1	PBW	-56	-0.0108	-7470A	PPB	PPB	16 Jan 2017 14:56:22	U	SMPL
PB95981BS - 1	LCSW	13462	3.6264	-7470A	PPB	PPB	16 Jan 2017 14:58:29	U	SMPL
PB95981TB - 1	PB95963TB	-26	-0.0027	-7470A	PPB	PPB	16 Jan 2017 15:00:37	U	SMPL
I1162-17 - 1	SB-03-COMP	0	0.0043	-7470A	PPB	PPB	16 Jan 2017 15:02:45	U	SMPL
I1162-18 - 1	SB-45-COMP	-2	0.0037	-7470A	PPB	PPB	16 Jan 2017 15:04:53	U	SMPL
I1162-19 - 1	SB-46-COMP	6	0.0059	-7470A	PPB	PPB	16 Jan 2017 15:07:02	U	SMPL
I1162-20 - 1	SB-47-COMP	-17	-0.0003	-7470A	PPB	PPB	16 Jan 2017 15:09:09	U	SMPL
CCV67 - 1	CCV67	18407	4.9569	-7470A	PPB	PPB	16 Jan 2017 15:11:16	U	SMPL
CCB67 - 1	CCB67	-35	-0.0051	-7470A	PPB	PPB	16 Jan 2017 15:13:22	U	SMPL
I1162-21 - 1	SB-48-COMP	-13	0.0008	-7470A	PPB	PPB	16 Jan 2017 15:15:29	U	SMPL
I1162-22 - 1	SB-49-COMP	-5	0.0029	-7470A	PPB	PPB	16 Jan 2017 15:17:36	U	SMPL
I1162-23 - 1	SB-08-COMP	-5	0.0029	-7470A	PPB	PPB	16 Jan 2017 15:19:43	U	SMPL
I1162-24 - 1	SB-09-COMP	1	0.0045	-7470A	PPB	PPB	16 Jan 2017 15:21:50	U	SMPL
I1162-24DUP - 1	SB-09-COMPDUP	6	0.0059	-7470A	PPB	PPB	16 Jan 2017 15:23:58	U	SMPL
I1162-24MS - 1	SB-09-COMPMS	12358	3.3293	-7470A	PPB	PPB	16 Jan 2017 15:26:05	U	SMPL
I1162-24MSD - 1	SB-09-COMPMSD	13908	3.7464	-7470A	PPB	PPB	16 Jan 2017 15:28:14	U	SMPL
I1162-24LX5 - 1		-32	-0.0043	-7470A	PPB	PPB	16 Jan 2017 15:30:22	U	SMPL
I1162-24A - 1		15679	4.2229	-7470A	PPB	PPB	16 Jan 2017 15:32:30	U	SMPL
CCV68 - 1	CCV68	18370	4.9469	-7470A	PPB	PPB	16 Jan 2017 15:34:37	U	SMPL
CCB68 - 1	CCB68	-40	-0.0065	-7470A	PPB	PPB	16 Jan 2017 15:36:44	U	SMPL