



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85365

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/16/2017	18:10
2	CAL2	1	Water			7.00	01/16/2017	18:11
3	CAL3	1	Water			10.04	01/16/2017	18:12
4	ICV	1	Water			7.03	01/16/2017	18:13
5	CCV1	1	Water			2.02	01/16/2017	18:15
6	I1246-03	1	Solid	20.02	20	8.77	01/16/2017	18:16
7	I1246-03DUP	1	Solid	20.04	20	8.78	01/16/2017	18:17
8	CCV2	1	Water			12.01	01/16/2017	18:18

1385365

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY 11246 WorkList ID : 94462 Date : 1/16/2017 5:41:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/19/2017	Solid	I1246-03	Corrosivity	Cool 4 deg C	PSEG01	D41	ROW-SP-0-2-COMP	01/16/2017	9045C

Date/Time 01-16-17 5:40 PM
Received by: HL
Relinquished by: CP

Date/Time 01-16-17 7:00 PM
Received by: CP
Relinquished by: JP